

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892  
 Lab File ID: VY020373.D Date Analyzed: 11/21/2024  
 Instrument ID: MSVOA\_Y Time Analyzed: 09:25  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 121826        | 13.352 |  |  |  |  |
| UPPER LIMIT    | 243652        | 13.852 |  |  |  |  |
| LOWER LIMIT    | 60913         | 12.852 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| WB-310-TOP     | 113213        | 13.35  |  |  |  |  |
| VY1121SBL01    | 89046         | 13.35  |  |  |  |  |
| VY1121SBS01    | 121310        | 13.35  |  |  |  |  |
| VY1121SBSD01   | 112550        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.