

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892
Lab File ID: VY020372.D BFB Injection Date: 11/21/2024
Instrument ID: MSVOA_Y BFB Injection Time: 08:32
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.5
75	30.0 - 60.0% of mass 95	55.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 100.0% of mass 95	73.7
175	5.0 - 9.0% of mass 174	5.8 (7.9) 1
176	95.0 - 101.0% of mass 174	70.4 (95.6) 1
177	5.0 - 9.0% of mass 176	4.7 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY020373.D	11/21/2024	09:25
VY1121SBL01	VY1121SBL01	VY020374.D	11/21/2024	09:59
VY1121SBS01	VY1121SBS01	VY020375.D	11/21/2024	10:39
VY1121SBSD01	VY1121SBSD01	VY020376.D	11/21/2024	11:02
WB-310-TOP	P4892-01	VY020395.D	11/21/2024	18:27