

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHB2
 Preparation Blank Matrix : Soil
 Preparation Blank Concentration Units (µg/L, mg/L, mg/kg dry weight, or µg): mg/kg
 Analytical Method: ICP-AES Preparation Batch: PB165105
 Run Batch: LB133566 Preparation Method: 3050B

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID: ICB007	Q	ID: CCB041	Q	ID: CCB042	Q	ID: CCB043	Q	ID: PBS105	Q
Aluminum	200	U	200	U	200	U	200	U	20.0	U
Antimony	60.0	U	60.0	U	60.0	U	60.0	U	6.0	U
Arsenic	10.0	U	-7.9	J	10.0	U	10.0	U	1.0	U
Barium	200	U	200	U	200	U	200	U	20.0	U
Beryllium	5.0	U	5.0	U	5.0	U	5.0	U	0.5	U
Cadmium	5.0	U	5.0	U	5.0	U	5.0	U	0.5	U
Calcium	5000	U	5000	U	5000	U	5000	U	500	U
Chromium	10.0	U	10.0	U	10.0	U	1.0	J	1.0	U
Cobalt	50.0	U	50.0	U	50.0	U	50.0	U	5.0	U
Copper	25.0	U	25.0	U	25.0	U	25.0	U	2.5	U
Iron	100	U	26	J	34	J	23	J	2.6	J
Lead	10.0	U	4.3	J	10.0	U	4.0	J	1.0	U
Magnesium	5000	U	5000	U	5000	U	5000	U	500	U
Manganese	15.0	U	15.0	U	15.0	U	15.0	U	1.5	U
Nickel	40.0	U	40.0	U	40.0	U	40.0	U	4.0	U
Potassium	5000	U	5000	U	5000	U	5000	U	500	U
Selenium	35.0	U	35.0	U	35.0	U	35.0	U	3.5	U
Silver	10.0	U	-1.2	J	-1.4	J	10.0	U	-0.2	J
Sodium	5000	U	5000	U	5000	U	5000	U	500	U
Thallium	25.0	U	25.0	U	25.0	U	25.0	U	2.5	U
Vanadium	50.0	U	50.0	U	50.0	U	50.0	U	5.0	U
Zinc	60.0	U	60.0	U	60.0	U	60.0	U	6.0	U

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHHB2

Preparation Blank Matrix :

Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg):

Analytical Method: ICP-AES Preparation Batch:

Run Batch: LB133566 Preparation Method:

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB044	Q	ID: CCB045	Q	ID: CCB046	Q	ID:	Q
Aluminum			24	J	22	J	200	U		
Antimony			60.0	U	60.0	U	60.0	U		
Arsenic			10.0	U	10.0	U	-5.2	J		
Barium			200	U	200	U	200	U		
Beryllium			5.0	U	5.0	U	5.0	U		
Cadmium			5.0	U	5.0	U	5.0	U		
Calcium			5000	U	5000	U	5000	U		
Chromium			10.0	U	10.0	U	10.0	U		
Cobalt			50.0	U	50.0	U	50.0	U		
Copper			25.0	U	25.0	U	25.0	U		
Iron			26	J	36	J	43	J		
Lead			10.0	U	3.8	J	3.5	J		
Magnesium			5000	U	5000	U	5000	U		
Manganese			15.0	U	15.0	U	15.0	U		
Nickel			40.0	U	40.0	U	40.0	U		
Potassium			5000	U	600	J	5000	U		
Selenium			35.0	U	35.0	U	35.0	U		
Silver			-2.2	J	10.0	U	10.0	U		
Sodium			5000	U	5000	U	5000	U		
Thallium			25.0	U	25.0	U	25.0	U		
Vanadium			50.0	U	50.0	U	50.0	U		
Zinc			60.0	U	60.0	U	60.0	U		

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHHB2

Preparation Blank Matrix :

Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg):

Analytical Method: ICP-AES Preparation Batch:

Run Batch: LB133566 Preparation Method:

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB047	Q	ID:	Q	ID:	Q	ID:	Q
Aluminum			29	J						
Antimony			60.0	U						
Arsenic			10.0	U						
Barium			200	U						
Beryllium			5.0	U						
Cadmium			5.0	U						
Calcium			5000	U						
Chromium			10.0	U						
Cobalt			50.0	U						
Copper			25.0	U						
Iron			37	J						
Lead			3.3	J						
Magnesium			5000	U						
Manganese			15.0	U						
Nickel			40.0	U						
Potassium			5000	U						
Selenium			35.0	U						
Silver			-1.6	J						
Sodium			5000	U						
Thallium			25.0	U						
Vanadium			50.0	U						
Zinc			60.0	U						

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHHB2

Preparation Blank Matrix :

Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg):

Analytical Method: ICP-AES Preparation Batch:

Run Batch: LB133698 Preparation Method:

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID: ICB011	Q	ID: CCB071	Q	ID: CCB072	Q	ID: CCB073	Q	ID:	Q
Aluminum	130	J	200	U	100	J	45	J		
Antimony	60.0	U	60.0	U	60.0	U	60.0	U		
Arsenic	5.6	J	6.1	J	10.0	U	8.7	J		
Barium	200	U	200	U	200	U	200	U		
Beryllium	5.0	U	5.0	U	5.0	U	5.0	U		
Cadmium	5.0	U	5.0	U	5.0	U	5.0	U		
Calcium	5000	U	5000	U	5000	U	5000	U		
Chromium	1.0	J	0.85	J	10.0	U	1.2	J		
Cobalt	50.0	U	50.0	U	50.0	U	50.0	U		
Copper	25.0	U	25.0	U	25.0	U	25.0	U		
Iron	100	U	100	U	100	U	100	U		
Lead	-3.3	J	-3.4	J	10.0	U	10.0	U		
Magnesium	5000	U	5000	U	5000	U	5000	U		
Manganese	15.0	U	15.0	U	15.0	U	15.0	U		
Nickel	40.0	U	40.0	U	40.0	U	40.0	U		
Potassium	5000	U	5000	U	5000	U	5000	U		
Selenium	35.0	U	35.0	U	35.0	U	35.0	U		
Silver	10.0	U	10.0	U	10.0	U	10.0	U		
Sodium	5000	U	5000	U	5000	U	5000	U		
Thallium	25.0	U	25.0	U	25.0	U	25.0	U		
Vanadium	50.0	U	50.0	U	50.0	U	50.0	U		
Zinc	60.0	U	60.0	U	60.0	U	60.0	U		

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHHB2

Preparation Blank Matrix :

Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg):

Analytical Method: ICP-AES Preparation Batch:

Run Batch: LB133698 Preparation Method:

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB074	Q	ID: CCB075	Q	ID: CCB076	Q	ID:	Q
Aluminum			200	U	52	J	200	U		
Antimony			60.0	U	60.0	U	60.0	U		
Arsenic			10.0	U	6.9	J	10.0	U		
Barium			200	U	200	U	200	U		
Beryllium			5.0	U	5.0	U	5.0	U		
Cadmium			5.0	U	5.0	U	5.0	U		
Calcium			5000	U	5000	U	5000	U		
Chromium			10.0	U	10.0	U	1.7	J		
Cobalt			50.0	U	50.0	U	50.0	U		
Copper			25.0	U	25.0	U	25.0	U		
Iron			100	U	100	U	22	J		
Lead			10.0	U	10.0	U	10.0	U		
Magnesium			5000	U	5000	U	5000	U		
Manganese			15.0	U	15.0	U	15.0	U		
Nickel			40.0	U	40.0	U	40.0	U		
Potassium			5000	U	5000	U	-620	J		
Selenium			35.0	U	35.0	U	35.0	U		
Silver			10.0	U	10.0	U	10.0	U		
Sodium			5000	U	5000	U	5000	U		
Thallium			25.0	U	25.0	U	25.0	U		
Vanadium			50.0	U	50.0	U	50.0	U		
Zinc			60.0	U	60.0	U	60.0	U		

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHHB2

Preparation Blank Matrix :

Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg):

Analytical Method: ICP-AES Preparation Batch:

Run Batch: LB133698 Preparation Method:

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB077	Q	ID: CCB078	Q	ID: CCB079	Q	ID:	Q
Aluminum			200	U	200	U	200	U		
Antimony			60.0	U	60.0	U	60.0	U		
Arsenic			7.8	J	10.0	U	7.6	J		
Barium			200	U	200	U	200	U		
Beryllium			5.0	U	5.0	U	5.0	U		
Cadmium			5.0	U	5.0	U	5.0	U		
Calcium			5000	U	5000	U	5000	U		
Chromium			10.0	U	1.1	J	1.0	J		
Cobalt			50.0	U	50.0	U	50.0	U		
Copper			25.0	U	25.0	U	25.0	U		
Iron			100	U	38	J	100	U		
Lead			10.0	U	10.0	U	10.0	U		
Magnesium			5000	U	5000	U	5000	U		
Manganese			15.0	U	1.7	J	15.0	U		
Nickel			40.0	U	40.0	U	40.0	U		
Potassium			-560	J	-680	J	-650	J		
Selenium			35.0	U	35.0	U	35.0	U		
Silver			10.0	U	10.0	U	10.0	U		
Sodium			5000	U	5000	U	5000	U		
Thallium			25.0	U	25.0	U	25.0	U		
Vanadium			50.0	U	-6.4	J	50.0	U		
Zinc			60.0	U	60.0	U	60.0	U		

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHHB2

Preparation Blank Matrix :

Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg):

Analytical Method: ICP-AES Preparation Batch:

Run Batch: LB133698 Preparation Method:

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB080	Q	ID: CCB081	Q	ID:	Q	ID:	Q
Aluminum			200	U	200	U				
Antimony			60.0	U	60.0	U				
Arsenic			6.9	J	10.0	U				
Barium			200	U	200	U				
Beryllium			5.0	U	5.0	U				
Cadmium			5.0	U	5.0	U				
Calcium			5000	U	5000	U				
Chromium			0.95	J	10.0	U				
Cobalt			50.0	U	50.0	U				
Copper			25.0	U	25.0	U				
Iron			20	J	100	U				
Lead			10.0	U	10.0	U				
Magnesium			5000	U	5000	U				
Manganese			15.0	U	15.0	U				
Nickel			40.0	U	40.0	U				
Potassium			-720	J	-750	J				
Selenium			35.0	U	35.0	U				
Silver			10.0	U	10.0	U				
Sodium			5000	U	5000	U				
Thallium			25.0	U	25.0	U				
Vanadium			50.0	U	50.0	U				
Zinc			60.0	U	60.0	U				