

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133566

Review By	Sarabjit Jaswal	Review On	11/22/2024 4:49:16 AM
Supervise By	Kareem Khairalla	Supervise On	12/5/2024 5:24:09 PM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83148,MP83149

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/21/24 11:55		Kareem	OK
2	S1	S01	CAL2	11/21/24 11:59		Kareem	OK
3	S2	S02	CAL3	11/21/24 12:04		Kareem	OK
4	S3	S03	CAL4	11/21/24 12:08		Kareem	OK
5	S4	S04	CAL5	11/21/24 12:13		Kareem	OK
6	S5	S05	CAL6	11/21/24 12:17		Kareem	OK
7	S6	S06	CAL7	11/21/24 12:22		Kareem	OK
8	ICV007	ICV007	ICV	11/21/24 12:27		Kareem	OK
9	ICB007	ICB007	ICB	11/21/24 12:33		Kareem	OK
10	ICSA007	ICSA007	ICSA	11/21/24 13:20	Inst. paused for review	Kareem	OK
11	ICSAB007	ICSAB007	ICSAB	11/21/24 13:24		Kareem	OK
12	CCV041	CCV041	CCV	11/21/24 13:29		Kareem	OK
13	CCB041	CCB041	CCB	11/21/24 13:33		Kareem	OK
14	P4907-01	MBHHE1	SAM	11/21/24 13:38		Kareem	OK
15	P4907-02	MBHHE1D	DUP	11/21/24 13:42		Kareem	OK
16	P4907-01L	MBHHE1L	SD	11/21/24 13:47		Kareem	OK
17	P4907-04	MBHHE4	SAM	11/21/24 13:56	Ca High	Kareem	Dilution
18	P4907-05	MBHHE5	SAM	11/21/24 14:01	Ca High	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133566

Review By	Sarabjit Jaswal	Review On	11/22/2024 4:49:16 AM
Supervise By	Kareem Khairalla	Supervise On	12/5/2024 5:24:09 PM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83148,MP83149

19	P4907-06	MBHHE6	SAM	11/21/24 14:05		Kareem	OK
20	P4907-07	MBHHE7	SAM	11/21/24 14:10		Kareem	OK
21	P4907-08	MBHHE8	SAM	11/21/24 14:14		Kareem	OK
22	P4907-09	MBHHE9	SAM	11/21/24 14:19		Kareem	OK
23	P4907-10	MBHHE9	SAM	11/21/24 14:23		Kareem	OK
24	P4907-11	MBHHC0	SAM	11/21/24 14:27		Kareem	OK
25	P4907-12	MBHHC1	SAM	11/21/24 14:32		Kareem	OK
26	P4907-13	MBHHC2	SAM	11/21/24 14:36		Kareem	OK
27	P4907-14	MBHHC3	SAM	11/21/24 14:41		Kareem	OK
28	P4907-15	MBHHC4	SAM	11/21/24 14:45		Kareem	OK
29	P4907-16	MBHHD6	SAM	11/21/24 14:50		Kareem	OK
30	P4907-17	MBHHD7	SAM	11/21/24 14:54		Kareem	OK
31	PB165102BL	PBS102	MB	11/21/24 14:59		Kareem	OK
32	PB165102BS	LCS102	LCS	11/21/24 15:03		Kareem	OK
33	CCV042	CCV042	CCV	11/21/24 15:08		Kareem	OK
34	CCB042	CCB042	CCB	11/21/24 15:12		Kareem	OK
35	P4906-01	MBHHE2	SAM	11/21/24 15:17		Kareem	OK
36	P4906-02	MBHHE3	SAM	11/21/24 15:21		Kareem	OK
37	P4906-03	MBHHE3D	DUP	11/21/24 15:26		Kareem	OK
38	P4906-02L	MBHHE3L	SD	11/21/24 15:30		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133566

Review By	Sarabjit Jaswal	Review On	11/22/2024 4:49:16 AM
Supervise By	Kareem Khairalla	Supervise On	12/5/2024 5:24:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

39	P4906-05	MBHHB4	SAM	11/21/24 15:39		Kareem	OK
40	P4906-06	MBHHB5	SAM	11/21/24 15:43		Kareem	OK
41	P4906-07	MBHHD8	SAM	11/21/24 15:48		Kareem	OK
42	P4906-08	MBHHD9	SAM	11/21/24 15:52		Kareem	OK
43	P4906-09	MBHHE0	SAM	11/21/24 15:57		Kareem	OK
44	P4906-11	MBHHE2	SAM	11/21/24 16:01		Kareem	OK
45	P4906-12	MBHHE3	SAM	11/21/24 16:05		Kareem	OK
46	P4906-13	MBHH98	SAM	11/21/24 16:10		Kareem	OK
47	P4906-14	MBHH99	SAM	11/21/24 16:14		Kareem	OK
48	P4906-15	MBHHA0	SAM	11/21/24 16:19		Kareem	OK
49	P4906-16	MBHHA1	SAM	11/21/24 16:23		Kareem	OK
50	P4906-17	MBHHA2	SAM	11/21/24 16:27		Kareem	OK
51	P4906-18	MBHHA3	SAM	11/21/24 16:32	Ca High	Kareem	Dilution
52	P4906-19	MBHHA4	SAM	11/21/24 16:36	Ca High	Kareem	Dilution
53	P4906-20	MBHHB6	SAM	11/21/24 16:41		Kareem	OK
54	CCV043	CCV043	CCV	11/21/24 16:45		Kareem	OK
55	CCB043	CCB043	CCB	11/21/24 16:50		Kareem	OK
56	P4906-21	MBHHB7	SAM	11/21/24 16:54		Kareem	OK
57	P4906-22	MBHHB8	SAM	11/21/24 16:59		Kareem	OK
58	PB165105BL	PBS105	MB	11/21/24 17:03		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133566

Review By	Sarabjit Jaswal	Review On	11/22/2024 4:49:16 AM
Supervise By	Kareem Khairalla	Supervise On	12/5/2024 5:24:09 PM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83148,MP83149

59	PB165105BS	LCS105	LCS	11/21/24 17:07		Kareem	OK
60	P4913-01	MBHHG2	SAM	11/21/24 17:12		Kareem	OK
61	P4913-02	MBHHG3	SAM	11/21/24 17:16		Kareem	OK
62	P4913-03	MBHHG4	SAM	11/21/24 17:21		Kareem	OK
63	P4913-04	MBHHG5	SAM	11/21/24 17:25		Kareem	OK
64	P4913-05	MBHHG6	SAM	11/21/24 17:30		Kareem	OK
65	P4913-06	MBHHG7	SAM	11/21/24 17:34		Kareem	OK
66	P4913-07	MBHHG8	SAM	11/21/24 17:38		Kareem	OK
67	P4913-08	MBHHG9	SAM	11/21/24 17:43		Kareem	OK
68	P4913-09	MBHHH0	SAM	11/21/24 17:52		Kareem	OK
69	P4913-10	MBHHH1	SAM	11/21/24 17:56		Kareem	OK
70	P4913-11	MBHHH2	SAM	11/21/24 18:05		Kareem	OK
71	P4913-12	MBHHH3	SAM	11/21/24 18:09		Kareem	OK
72	P4913-13	MBHHH4	SAM	11/21/24 18:14		Kareem	OK
73	P4913-14	MBHHH5	SAM	11/21/24 18:18		Kareem	OK
74	P4913-15	MBHHH6	SAM	11/21/24 18:23		Kareem	OK
75	P4913-16	MBHHH6D	DUP	11/21/24 18:27		Kareem	OK
76	CCV044	CCV044	CCV	11/21/24 18:36		Kareem	OK
77	CCB044	CCB044	CCB	11/21/24 18:40		Kareem	OK
78	P4906-04	MBHHB3S	MS	11/21/24 18:45	MS Fail For Zn	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133566

Review By	Sarabjit Jaswal	Review On	11/22/2024 4:49:16 AM
Supervise By	Kareem Khairalla	Supervise On	12/5/2024 5:24:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

79	P4913-15L	MBHHH6L	SD	11/21/24 18:49		Kareem	OK
80	P4907-03	MBHHE1S	MS	11/21/24 18:54	MS Fail For Be(1862.50ppb),Se(Below RL),	Kareem	OK
81	P4913-17	MBHHH6S	MS	11/21/24 19:07	MS Fail For Se(Below RL)	Kareem	OK
82	P4912-04	MBHHJ4	SAM	11/21/24 19:11	Not Use	Kareem	Not Ok
83	P4912-06	MBHHJ4S	MS	11/21/24 19:16	Not Use	Kareem	Not Ok
84	P4912-04L	MBHHJ4L	SD	11/21/24 19:20	Not Use	Kareem	Not Ok
85	P4912-05	MBHHJ4D	DUP	11/21/24 19:25	Not Use	Kareem	Not Ok
86	P4913-18	MBHHH7	SAM	11/21/24 19:29		Kareem	OK
87	P4913-19	MBHHH8	SAM	11/21/24 19:33		Kareem	OK
88	P4913-20	MBHHH9	SAM	11/21/24 19:38		Kareem	OK
89	P4913-21	MBHHJ0	SAM	11/21/24 19:42		Kareem	OK
90	PB165145BL	PBS145	MB	11/21/24 19:47		Kareem	OK
91	PB165145BS	LCS145	LCS	11/21/24 19:51		Kareem	OK
92	P4912-01	MBHHJ1	SAM	11/21/24 19:56	Not Use	Kareem	Not Ok
93	P4912-02	MBHHJ2	SAM	11/21/24 20:00	Not Use	Kareem	Not Ok
94	P4912-03	MBHHJ3	SAM	11/21/24 20:04	Not Use	Kareem	Not Ok
95	P4912-07	MBHHJ5	SAM	11/21/24 20:09	Not Use	Kareem	Not Ok
96	P4912-08	MBHHJ6	SAM	11/21/24 20:13	Not Use	Kareem	Not Ok
97	P4912-09	MBHHJ7	SAM	11/21/24 20:18	Not Use	Kareem	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133566

Review By	Sarabjit Jaswal	Review On	11/22/2024 4:49:16 AM
Supervise By	Kareem Khairalla	Supervise On	12/5/2024 5:24:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

98	CCV045	CCV045	CCV	11/21/24 20:22		Kareem	OK
99	CCB045	CCB045	CCB	11/21/24 20:26		Kareem	OK
100	P4912-10	MBHHJ8	SAM	11/21/24 20:31	Not Use	Kareem	Not Ok
101	P4912-11	MBHHJ9	SAM	11/21/24 20:35	Not Use	Kareem	Not Ok
102	P4912-12	MBHHK0	SAM	11/21/24 20:40	Not Use	Kareem	Not Ok
103	P4912-13	MBHHK1	SAM	11/21/24 20:44	Not Use	Kareem	Not Ok
104	P4912-14	MBHHK2	SAM	11/21/24 20:48	Not Use	Kareem	Not Ok
105	P4912-15	MBHHK3	SAM	11/21/24 20:53	Not Use	Kareem	Not Ok
106	P4912-16	MBHHL8	SAM	11/21/24 20:57	Not Use	Kareem	Not Ok
107	P4912-17	MBHHL9	SAM	11/21/24 21:02	Not Use	Kareem	Not Ok
108	P4912-18	MBHBM0	SAM	11/21/24 21:06	Not Use	Kareem	Not Ok
109	P4912-19	MBHBM1	SAM	11/21/24 21:10	Not Use	Kareem	Not Ok
110	P4912-20	MBHBM2	SAM	11/21/24 21:15	Not Use	Kareem	Not Ok
111	P4912-21	MBHBM3	SAM	11/21/24 21:19	Not Use	Kareem	Not Ok
112	P4912-22	MBHBM4	SAM	11/21/24 21:23	Not Use	Kareem	Not Ok
113	PB165139BL	PBS139	MB	11/21/24 21:28	Not Use	Kareem	Not Ok
114	CCV046	CCV046	CCV	11/21/24 21:32		Kareem	OK
115	CCB046	CCB046	CCB	11/21/24 21:37		Kareem	OK
116	PB165139BS	LCS139	LCS	11/21/24 21:41	Not Use	Kareem	Not Ok
117	P4907-04DL	MBHHE4	SAM	11/21/24 21:46	Not Use	Kareem	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133566

Review By	Sarabjit Jaswal	Review On	11/22/2024 4:49:16 AM
Supervise By	Kareem Khairalla	Supervise On	12/5/2024 5:24:09 PM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83148,MP83149

118	P4907-05DL	MBHHE5	SAM	11/21/24 21:50	Not Use	Kareem	Not Ok
119	P4906-18DL	MBHHA3	SAM	11/21/24 21:55	Not Use	Kareem	Not Ok
120	P4906-19DL	MBHHA4	SAM	11/21/24 22:00	Not Use	Kareem	Not Ok
121	P4906-02A	MBHHB3A	PS	11/21/24 22:04	PS For Zn - 0.03ml m5819-10ml sample	Kareem	OK
122	P4907-01A	MBHHE1A	PS	11/21/24 22:08	Not use	Kareem	Not Ok
123	P4913-15A	MBHHH6A	PS	11/21/24 22:13	PS For Se - 0.07 - MP83152 - 10 ml Sample	Kareem	OK
124	CCV047	CCV047	CCV	11/21/24 22:17	%RSD Fail For Al,Ba,Be,Ca,Cu,Fe,Mg,Mn, K,Na,V	Kareem	OK
125	CCB047	CCB047	CCB	11/21/24 22:22		Kareem	OK