

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHJ1
 Analytical Method: ICP-AES ICSA Source: EPA-1211
 Instrument ID: P5 ICSB Source: EPA-0710
 Run Batch: LB133698
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	240000	94	240000	99
Antimony	0	618	3.4		560	90
Arsenic	0	104	8.1		110	106
Barium	6	537	4.7	77	530	99
Beryllium	0	495	0.1		530	106
Cadmium	1	972	2.5	247	1100	109
Calcium	245000	235000	250000	100	250000	106
Chromium	52	542	50	96	610	113
Cobalt	0	476	2.0		520	110
Copper	2	511	-1.8	-90	480	94
Iron	101000	99300	100000	100	100000	103
Lead	0	49	-3.6		45	91
Magnesium	255000	248000	240000	93	240000	97
Manganese	7	507	14	192	530	105
Nickel	2	954	4.0	199	1100	110
Potassium	0	0	-850		-970	
Selenium	0	46	-2.2		43	93
Silver	0	201	3.5		190	95
Sodium	0	0	640		720	
Thallium	0	108	-9.9		96	89
Vanadium	0	491	-3.6		540	109
Zinc	0	952	12		920	97

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHJ1
 Analytical Method: ICP-AES ICSA Source: EPA-1211
 Instrument ID: P5 ICSB Source: EPA-0710
 Run Batch: LB133774
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	230000	89	240000	99
Antimony	0	618	5.9		600	96
Arsenic	0	104	4.2		95	91
Barium	6	537	5.6	93	530	99
Beryllium	0	495	0.87		530	107
Cadmium	1	972	-0.46	-46	1000	103
Calcium	245000	235000	240000	97	260000	109
Chromium	52	542	51	97	610	113
Cobalt	0	476	-1.4		520	109
Copper	2	511	-5.0	-251	480	95
Iron	101000	99300	96000	95	100000	103
Lead	0	49	-5.4		49	99
Magnesium	255000	248000	230000	89	240000	97
Manganese	7	507	12	165	540	106
Nickel	2	954	2.3	116	1000	109
Potassium	0	0	-300		-550	
Selenium	0	46	-0.36		43	93
Silver	0	201	-7.0		180	91
Sodium	0	0	620		620	
Thallium	0	108	-13		96	89
Vanadium	0	491	4.3		510	105
Zinc	0	952	-36		960	100

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHJ1
 Analytical Method: ICP-AES ICSA Source: EPA-1211
 Instrument ID: P5 ICSB Source: EPA-0710
 Run Batch: LB133883
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	240000	95	250000	101
Antimony	0	618	-6.8		560	90
Arsenic	0	104	-9.3		89	86
Barium	6	537	7.3	122	550	102
Beryllium	0	495	1.4		550	110
Cadmium	1	972	0.85	85	1000	104
Calcium	245000	235000	250000	102	260000	109
Chromium	52	542	52	100	620	115
Cobalt	0	476	32		540	114
Copper	2	511	-7.0	-350	490	95
Iron	101000	99300	100000	102	110000	106
Lead	0	49	-9.9		42	87
Magnesium	255000	248000	260000	100	260000	106
Manganese	7	507	10	146	550	108
Nickel	2	954	5.1	256	1100	113
Potassium	0	0	-290		-600	
Selenium	0	46	-16		27	58
Silver	0	201	-7.6		190	94
Sodium	0	0	-51		-94	
Thallium	0	108	7.7		110	101
Vanadium	0	491	0.48		530	108
Zinc	0	952	-36		1100	110