

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51879 MA No.: \_\_\_\_\_ SDG No.: MBHHF2  
 Instrument ID: P5 Start Date: 12/02/2024  
 Analytical Method: ICP-AES Run Batch: LB133698  
 Concentration Units: ug/L

| Analyte   | True | Found  | %D | True | Found | %D  | True  | Found | %D  |
|-----------|------|--------|----|------|-------|-----|-------|-------|-----|
| Aluminum  | 0    | 0.010  | 0  | 200  | 200   | 0   | 38750 | 38000 | -2  |
| Antimony  | 0    | 0.010  | 0  | 60   | 57    | -5  | 1250  | 1100  | -12 |
| Arsenic   | 0    | -0     | 0  | 10   | 12    | 20  | 1250  | 1100  | -12 |
| Barium    | 0    | -0.030 | 0  | 200  | 230   | 15  | 2500  | 2500  | 0   |
| Beryllium | 0    | -0     | 0  | 5    | 6.1   | 22  | 125   | 130   | 4   |
| Cadmium   | 0    | -0     | 0  | 5    | 5.6   | 12  | 625   | 620   | -1  |
| Calcium   | 0    | -0.97  | 0  | 5000 | 5800  | 16  | 62500 | 64000 | 2   |
| Chromium  | 0    | -0     | 0  | 10   | 13    | 30  | 3875  | 4100  | 6   |
| Cobalt    | 0    | -0     | 0  | 50   | 52    | 4   | 625   | 610   | -2  |
| Copper    | 0    | -0     | 0  | 25   | 27    | 8   | 3750  | 3800  | 1   |
| Iron      | 0    | -0.020 | 0  | 100  | 110   | 10  | 37500 | 39000 | 4   |
| Lead      | 0    | 0.0    | 0  | 10   | 8.1   | -19 | 6250  | 6200  | -1  |
| Magnesium | 0    | -0.060 | 0  | 5000 | 5400  | 8   | 62500 | 61000 | -2  |
| Manganese | 0    | -0     | 0  | 15   | 19    | 27  | 3750  | 3900  | 4   |
| Nickel    | 0    | -0.010 | 0  | 40   | 45    | 13  | 625   | 620   | -1  |
| Potassium | 0    | 0.30   | 0  | 5000 | 5100  | 2   | 37500 | 36000 | -4  |
| Selenium  | 0    | 0.0    | 0  | 35   | 37    | 6   | 1250  | 1100  | -12 |
| Silver    | 0    | -0     | 0  | 10   | 12    | 20  | 312.5 | 300   | -4  |
| Sodium    | 0    | -0.81  | 0  | 5000 | 5900  | 18  | 37500 | 38000 | 1   |
| Thallium  | 0    | -0.010 | 0  | 25   | 30    | 20  | 1250  | 1300  | 4   |
| Vanadium  | 0    | -0.010 | 0  | 50   | 60    | 20  | 625   | 640   | 2   |
| Zinc      | 0    | -0     | 0  | 60   | 71    | 18  | 3750  | 3400  | -9  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51879 MA No.: \_\_\_\_\_ SDG No.: MBHHF2  
 Instrument ID: P5 Start Date: 12/02/2024  
 Analytical Method: ICP-AES Run Batch: LB133698  
 Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True   | Found  | %D | True   | Found  | %D |
|-----------|--------|--------|----|--------|--------|----|--------|--------|----|
| Aluminum  | 77500  | 75000  | -3 | 155000 | 150000 | -3 | 310000 | 310000 | 0  |
| Antimony  | 2500   | 2300   | -8 | 5000   | 4700   | -6 | 10000  | 11000  | 10 |
| Arsenic   | 2500   | 2300   | -8 | 5000   | 4900   | -2 | 10000  | 10000  | 0  |
| Barium    | 5000   | 5000   | 0  | 10000  | 9900   | -1 | 20000  | 20000  | 0  |
| Beryllium | 250    | 250    | 0  | 500    | 500    | 0  | 1000   | 990    | -1 |
| Cadmium   | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Calcium   | 125000 | 130000 | 4  | 250000 | 250000 | 0  | 500000 | 510000 | 2  |
| Chromium  | 7750   | 8000   | 3  | 15500  | 16000  | 3  | 31000  | 30000  | -3 |
| Cobalt    | 1250   | 1200   | -4 | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Copper    | 7500   | 7500   | 0  | 15000  | 15000  | 0  | 30000  | 30000  | 0  |
| Iron      | 75000  | 77000  | 3  | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Lead      | 12500  | 13000  | 4  | 25000  | 25000  | 0  | 50000  | 50000  | 0  |
| Magnesium | 125000 | 120000 | -4 | 250000 | 240000 | -4 | 500000 | 500000 | 0  |
| Manganese | 7500   | 7600   | 1  | 15000  | 15000  | 0  | 30000  | 30000  | 0  |
| Nickel    | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Potassium | 75000  | 73000  | -3 | 150000 | 150000 | 0  | 300000 | 310000 | 3  |
| Selenium  | 2500   | 2300   | -8 | 5000   | 4700   | -6 | 10000  | 11000  | 10 |
| Silver    | 625    | 610    | -2 | 1250   | 1200   | -4 | 2500   | 2500   | 0  |
| Sodium    | 75000  | 74000  | -1 | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Thallium  | 2500   | 2600   | 4  | 5000   | 5000   | 0  | 10000  | 9900   | -1 |
| Vanadium  | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Zinc      | 7500   | 7000   | -7 | 15000  | 14000  | -7 | 30000  | 32000  | 7  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51879 MA No.:                      SDG No.: MBHHF2  
Instrument ID: P5 Start Date: 12/02/2024  
Analytical Method: ICP-AES Run Batch: LB133698  
Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True | Found | %D | True | Found | %D |
|-----------|--------|--------|----|------|-------|----|------|-------|----|
| Aluminum  | 800000 | 800000 | 0  |      |       |    |      |       |    |
| Calcium   | 800000 | 790000 | -1 |      |       |    |      |       |    |
| Iron      | 800000 | 800000 | 0  |      |       |    |      |       |    |
| Magnesium | 800000 | 810000 | 1  |      |       |    |      |       |    |
| Sodium    | 800000 | 800000 | 0  |      |       |    |      |       |    |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51879 MA No.: \_\_\_\_\_ SDG No.: MBHHF2  
 Instrument ID: P5 Start Date: 12/10/2024  
 Analytical Method: ICP-AES Run Batch: LB133883  
 Concentration Units: ug/L

| Analyte   | True | Found  | %D | True | Found | %D  | True  | Found | %D |
|-----------|------|--------|----|------|-------|-----|-------|-------|----|
| Aluminum  | 0    | -0     | 0  | 200  | 210   | 5   | 38750 | 38000 | -2 |
| Antimony  | 0    | 0.010  | 0  | 60   | 55    | -8  | 1250  | 1200  | -4 |
| Arsenic   | 0    | -0     | 0  | 10   | 11    | 10  | 1250  | 1400  | 12 |
| Barium    | 0    | -0.010 | 0  | 200  | 210   | 5   | 2500  | 2500  | 0  |
| Beryllium | 0    | -0     | 0  | 5    | 5.3   | 6   | 125   | 130   | 4  |
| Cadmium   | 0    | 0.0    | 0  | 5    | 4.7   | -6  | 625   | 630   | 1  |
| Calcium   | 0    | -0.50  | 0  | 5000 | 5200  | 4   | 62500 | 65000 | 4  |
| Chromium  | 0    | -0     | 0  | 10   | 10    | 0   | 3875  | 4000  | 3  |
| Cobalt    | 0    | -0     | 0  | 50   | 51    | 2   | 625   | 640   | 2  |
| Copper    | 0    | -0     | 0  | 25   | 28    | 12  | 3750  | 3800  | 1  |
| Iron      | 0    | -0.040 | 0  | 100  | 130   | 30  | 37500 | 40000 | 7  |
| Lead      | 0    | 0.0    | 0  | 10   | 9.0   | -10 | 6250  | 6200  | -1 |
| Magnesium | 0    | -0.83  | 0  | 5000 | 5100  | 2   | 62500 | 66000 | 6  |
| Manganese | 0    | -0     | 0  | 15   | 16    | 7   | 3750  | 3900  | 4  |
| Nickel    | 0    | 0.0    | 0  | 40   | 40    | 0   | 625   | 630   | 1  |
| Potassium | 0    | 0.38   | 0  | 5000 | 4800  | -4  | 37500 | 37000 | -1 |
| Selenium  | 0    | 0.0    | 0  | 35   | 36    | 3   | 1250  | 1200  | -4 |
| Silver    | 0    | 0.0    | 0  | 10   | 9.8   | -2  | 312.5 | 310   | -1 |
| Sodium    | 0    | -0.14  | 0  | 5000 | 5100  | 2   | 37500 | 38000 | 1  |
| Thallium  | 0    | -0     | 0  | 25   | 26    | 4   | 1250  | 1300  | 4  |
| Vanadium  | 0    | -0     | 0  | 50   | 53    | 6   | 625   | 630   | 1  |
| Zinc      | 0    | 0.010  | 0  | 60   | 55    | -8  | 3750  | 3400  | -9 |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51879 MA No.: \_\_\_\_\_ SDG No.: MBHHF2  
 Instrument ID: P5 Start Date: 12/10/2024  
 Analytical Method: ICP-AES Run Batch: LB133883  
 Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True   | Found  | %D | True   | Found  | %D |
|-----------|--------|--------|----|--------|--------|----|--------|--------|----|
| Aluminum  | 77500  | 76000  | -2 | 155000 | 150000 | -3 | 310000 | 310000 | 0  |
| Antimony  | 2500   | 2300   | -8 | 5000   | 4800   | -4 | 10000  | 10000  | 0  |
| Arsenic   | 2500   | 2600   | 4  | 5000   | 5000   | 0  | 10000  | 9700   | -3 |
| Barium    | 5000   | 5000   | 0  | 10000  | 10000  | 0  | 20000  | 20000  | 0  |
| Beryllium | 250    | 250    | 0  | 500    | 510    | 2  | 1000   | 990    | -1 |
| Cadmium   | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Calcium   | 125000 | 130000 | 4  | 250000 | 250000 | 0  | 500000 | 500000 | 0  |
| Chromium  | 7750   | 7900   | 2  | 15500  | 16000  | 3  | 31000  | 31000  | 0  |
| Cobalt    | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Copper    | 7500   | 7600   | 1  | 15000  | 15000  | 0  | 30000  | 30000  | 0  |
| Iron      | 75000  | 78000  | 4  | 150000 | 160000 | 7  | 300000 | 300000 | 0  |
| Lead      | 12500  | 12000  | -4 | 25000  | 25000  | 0  | 50000  | 50000  | 0  |
| Magnesium | 125000 | 130000 | 4  | 250000 | 260000 | 4  | 500000 | 510000 | 2  |
| Manganese | 7500   | 7700   | 3  | 15000  | 15000  | 0  | 30000  | 30000  | 0  |
| Nickel    | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Potassium | 75000  | 73000  | -3 | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Selenium  | 2500   | 2400   | -4 | 5000   | 4800   | -4 | 10000  | 10000  | 0  |
| Silver    | 625    | 620    | -1 | 1250   | 1200   | -4 | 2500   | 2500   | 0  |
| Sodium    | 75000  | 75000  | 0  | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Thallium  | 2500   | 2500   | 0  | 5000   | 5000   | 0  | 10000  | 9900   | -1 |
| Vanadium  | 1250   | 1200   | -4 | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Zinc      | 7500   | 6900   | -8 | 15000  | 14000  | -7 | 30000  | 32000  | 7  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51879 MA No.:                      SDG No.: MBHHF2  
Instrument ID: P5 Start Date: 12/10/2024  
Analytical Method: ICP-AES Run Batch: LB133883  
Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True | Found | %D | True | Found | %D |
|-----------|--------|--------|----|------|-------|----|------|-------|----|
| Aluminum  | 800000 | 810000 | 1  |      |       |    |      |       |    |
| Calcium   | 800000 | 800000 | 0  |      |       |    |      |       |    |
| Iron      | 800000 | 790000 | -1 |      |       |    |      |       |    |
| Magnesium | 800000 | 770000 | -4 |      |       |    |      |       |    |
| Sodium    | 800000 | 800000 | 0  |      |       |    |      |       |    |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51879 MA No.: \_\_\_\_\_ SDG No.: MBHHF2  
 Instrument ID: P5 Start Date: 12/13/2024  
 Analytical Method: ICP-AES Run Batch: LB133949  
 Concentration Units: ug/L

| Analyte   | True | Found | %D | True | Found | %D  | True  | Found | %D  |
|-----------|------|-------|----|------|-------|-----|-------|-------|-----|
| Aluminum  | 0    | 0.040 | 0  | 200  | 180   | -10 | 38750 | 37000 | -5  |
| Antimony  | 0    | 0.020 | 0  | 60   | 52    | -13 | 1250  | 1100  | -12 |
| Arsenic   | 0    | -0    | 0  | 10   | 13    | 30  | 1250  | 1300  | 4   |
| Barium    | 0    | 0.0   | 0  | 200  | 200   | 0   | 2500  | 2400  | -4  |
| Beryllium | 0    | -0    | 0  | 5    | 5.2   | 4   | 125   | 130   | 4   |
| Cadmium   | 0    | 0.0   | 0  | 5    | 4.8   | -4  | 625   | 610   | -2  |
| Calcium   | 0    | 0.32  | 0  | 5000 | 5000  | 0   | 62500 | 61000 | -2  |
| Chromium  | 0    | -0    | 0  | 10   | 9.9   | -1  | 3875  | 4000  | 3   |
| Cobalt    | 0    | -0    | 0  | 50   | 51    | 2   | 625   | 630   | 1   |
| Copper    | 0    | -0    | 0  | 25   | 29    | 16  | 3750  | 3700  | -1  |
| Iron      | 0    | 0.010 | 0  | 100  | 88    | -12 | 37500 | 38000 | 1   |
| Lead      | 0    | 0.0   | 0  | 10   | 8.3   | -17 | 6250  | 6000  | -4  |
| Magnesium | 0    | -0.37 | 0  | 5000 | 5300  | 6   | 62500 | 63000 | 1   |
| Manganese | 0    | -0    | 0  | 15   | 17    | 13  | 3750  | 3800  | 1   |
| Nickel    | 0    | 0.0   | 0  | 40   | 39    | -3  | 625   | 610   | -2  |
| Potassium | 0    | 0.30  | 0  | 5000 | 4900  | -2  | 37500 | 36000 | -4  |
| Selenium  | 0    | 0.010 | 0  | 35   | 26    | -26 | 1250  | 1100  | -12 |
| Silver    | 0    | 0.0   | 0  | 10   | 9.6   | -4  | 312.5 | 300   | -4  |
| Sodium    | 0    | -1.0  | 0  | 5000 | 6400  | 28  | 37500 | 37000 | -1  |
| Thallium  | 0    | -0    | 0  | 25   | 26    | 4   | 1250  | 1200  | -4  |
| Vanadium  | 0    | -0    | 0  | 50   | 52    | 4   | 625   | 610   | -2  |
| Zinc      | 0    | 0.020 | 0  | 60   | 50    | -17 | 3750  | 3400  | -9  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51879 MA No.: \_\_\_\_\_ SDG No.: MBHHF2  
 Instrument ID: P5 Start Date: 12/13/2024  
 Analytical Method: ICP-AES Run Batch: LB133949  
 Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True   | Found  | %D | True   | Found  | %D |
|-----------|--------|--------|----|--------|--------|----|--------|--------|----|
| Aluminum  | 77500  | 75000  | -3 | 155000 | 150000 | -3 | 310000 | 300000 | -3 |
| Antimony  | 2500   | 2300   | -8 | 5000   | 4800   | -4 | 10000  | 11000  | 10 |
| Arsenic   | 2500   | 2600   | 4  | 5000   | 5100   | 2  | 10000  | 9800   | -2 |
| Barium    | 5000   | 5000   | 0  | 10000  | 10000  | 0  | 20000  | 20000  | 0  |
| Beryllium | 250    | 260    | 4  | 500    | 510    | 2  | 1000   | 990    | -1 |
| Cadmium   | 1250   | 1200   | -4 | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Calcium   | 125000 | 120000 | -4 | 250000 | 250000 | 0  | 500000 | 490000 | -2 |
| Chromium  | 7750   | 7900   | 2  | 15500  | 16000  | 3  | 31000  | 30000  | -3 |
| Cobalt    | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Copper    | 7500   | 7600   | 1  | 15000  | 15000  | 0  | 30000  | 30000  | 0  |
| Iron      | 75000  | 77000  | 3  | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Lead      | 12500  | 12000  | -4 | 25000  | 25000  | 0  | 50000  | 50000  | 0  |
| Magnesium | 125000 | 130000 | 4  | 250000 | 250000 | 0  | 500000 | 490000 | -2 |
| Manganese | 7500   | 7700   | 3  | 15000  | 15000  | 0  | 30000  | 30000  | 0  |
| Nickel    | 1250   | 1200   | -4 | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Potassium | 75000  | 74000  | -1 | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Selenium  | 2500   | 2300   | -8 | 5000   | 4800   | -4 | 10000  | 11000  | 10 |
| Silver    | 625    | 620    | -1 | 1250   | 1200   | -4 | 2500   | 2500   | 0  |
| Sodium    | 75000  | 74000  | -1 | 150000 | 150000 | 0  | 300000 | 290000 | -3 |
| Thallium  | 2500   | 2500   | 0  | 5000   | 5000   | 0  | 10000  | 10000  | 0  |
| Vanadium  | 1250   | 1200   | -4 | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Zinc      | 7500   | 7000   | -7 | 15000  | 15000  | 0  | 30000  | 31000  | 3  |



FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51879 MA No.:                      SDG No.: MBHHF2  
Instrument ID: P5 Start Date: 12/13/2024  
Analytical Method: ICP-AES Run Batch: LB133949  
Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True | Found | %D | True | Found | %D |
|-----------|--------|--------|----|------|-------|----|------|-------|----|
| Aluminum  | 800000 | 820000 | 3  |      |       |    |      |       |    |
| Calcium   | 800000 | 820000 | 3  |      |       |    |      |       |    |
| Iron      | 800000 | 800000 | 0  |      |       |    |      |       |    |
| Magnesium | 800000 | 800000 | 0  |      |       |    |      |       |    |
| Sodium    | 800000 | 810000 | 1  |      |       |    |      |       |    |