

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133698 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV011				ID: CCV071				ID: CCV072		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	96	1	400000	420000	105	0	390000	98	1
Antimony	1000	940	94	0	5000	5100	102	1	4600	93	1
Arsenic	1000	930	93	0	5000	5400	107	0	5400	109	1
Barium	520	530	102	1	10000	11000	106	1	10000	100	1
Beryllium	510	540	105	1	500	520	104	0	490	98	1
Cadmium	510	510	99	0	2500	2700	107	0	2600	106	1
Calcium	10000	10000	103	1	400000	430000	107	0	400000	100	1
Chromium	520	570	110	0	15500	16000	105	0	15000	97	4
Cobalt	520	500	96	1	2500	2700	106	0	2500	100	1
Copper	510	500	98	1	15000	16000	105	0	15000	97	1
Iron	10000	11000	107	1	400000	430000	108	0	400000	99	1
Lead	1000	1000	100	0	25000	26000	105	0	25000	100	1
Magnesium	6000	5500	92	2	400000	420000	105	0	410000	103	1
Manganese	520	540	104	0	15000	16000	104	0	15000	98	1
Nickel	530	510	97	0	2500	2600	105	0	2500	99	1
Potassium	9900	9600	97	0	150000	160000	108	0	150000	100	1
Selenium	1000	980	98	0	5000	5000	100	1	4800	96	1
Silver	250	250	99	1	1250	1300	107	0	1300	100	4
Sodium	10000	10000	103	1	400000	420000	106	0	390000	98	1
Thallium	1000	1100	108	0	5000	5000	100	0	4900	98	1
Vanadium	500	540	108	1	2500	2600	105	0	2400	98	1
Zinc	1000	960	96	1	15000	16000	104	1	14000	94	4

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133698 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV073				ID: CCV074		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	1	400000	100	0
Antimony					5000	4700	93	2	4800	95	0
Arsenic					5000	5400	108	2	5400	108	0
Barium					10000	10000	100	2	10000	100	1
Beryllium					500	500	99	1	510	101	1
Cadmium					2500	2600	105	2	2600	105	0
Calcium					400000	400000	101	1	400000	101	0
Chromium					15500	15000	99	1	15000	99	0
Cobalt					2500	2500	100	2	2500	100	0
Copper					15000	15000	98	1	15000	99	0
Iron					400000	400000	101	1	400000	101	0
Lead					25000	25000	99	2	25000	100	0
Magnesium					400000	420000	104	1	420000	105	0
Manganese					15000	15000	98	1	15000	98	0
Nickel					2500	2500	99	2	2500	99	0
Potassium					150000	150000	101	1	150000	101	0
Selenium					5000	4900	97	2	5000	99	0
Silver					1250	1300	101	1	1300	101	0
Sodium					400000	400000	100	1	400000	100	0
Thallium					5000	5000	99	2	5100	102	0
Vanadium					2500	2500	99	1	2500	99	0
Zinc					15000	14000	95	1	14000	96	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHMF2Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133698 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV075				ID: CCV076		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	100	1	390000	98	1
Antimony					5000	4700	94	0	4700	94	0
Arsenic					5000	5400	107	0	5300	106	0
Barium					10000	10000	100	1	9800	98	1
Beryllium					500	520	103	1	520	103	0
Cadmium					2500	2600	104	0	2600	104	0
Calcium					400000	400000	101	1	400000	99	1
Chromium					15500	15000	99	0	15000	100	1
Cobalt					2500	2500	100	0	2500	99	0
Copper					15000	15000	99	1	15000	98	1
Iron					400000	410000	101	1	400000	101	1
Lead					25000	25000	99	0	24000	98	0
Magnesium					400000	420000	105	1	420000	104	1
Manganese					15000	15000	98	1	14000	96	1
Nickel					2500	2500	98	0	2400	97	0
Potassium					150000	150000	101	1	150000	100	1
Selenium					5000	4900	99	0	4900	98	0
Silver					1250	1200	100	0	1300	101	1
Sodium					400000	400000	100	1	400000	100	1
Thallium					5000	5100	102	0	5100	102	0
Vanadium					2500	2500	99	1	2500	99	1
Zinc					15000	14000	95	0	14000	96	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133698 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV077				ID: CCV078		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	1	400000	100	1
Antimony					5000	4700	95	1	4700	94	1
Arsenic					5000	5300	106	1	5200	105	1
Barium					10000	9700	97	1	9800	98	2
Beryllium					500	520	104	1	540	108	2
Cadmium					2500	2600	104	1	2600	102	1
Calcium					400000	400000	99	1	400000	100	2
Chromium					15500	15000	96	5	15000	98	7
Cobalt					2500	2500	99	1	2400	98	1
Copper					15000	15000	98	1	15000	100	2
Iron					400000	400000	101	1	410000	104	2
Lead					25000	24000	98	1	24000	96	1
Magnesium					400000	420000	105	1	400000	100	2
Manganese					15000	14000	96	1	15000	97	2
Nickel					2500	2400	98	1	2400	96	1
Potassium					150000	150000	100	1	150000	101	2
Selenium					5000	5000	100	1	5000	99	1
Silver					1250	1200	96	6	1200	98	8
Sodium					400000	400000	100	1	410000	102	2
Thallium					5000	5200	104	1	5200	104	1
Vanadium					2500	2500	99	0	2500	101	2
Zinc					15000	14000	91	6	14000	94	9

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHMF2Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133698 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV079				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	390000	99	1			
Antimony					5000	4700	94	0			
Arsenic					5000	5200	104	1			
Barium					10000	9500	95	1			
Beryllium					500	530	106	1			
Cadmium					2500	2600	102	0			
Calcium					400000	390000	98	1			
Chromium					15500	15000	100	0			
Cobalt					2500	2400	98	0			
Copper					15000	15000	97	1			
Iron					400000	400000	101	1			
Lead					25000	24000	96	0			
Magnesium					400000	390000	97	1			
Manganese					15000	14000	94	1			
Nickel					2500	2400	96	0			
Potassium					150000	150000	99	1			
Selenium					5000	5000	100	1			
Silver					1250	1200	100	0			
Sodium					400000	400000	100	1			
Thallium					5000	5200	104	1			
Vanadium					2500	2500	99	1			
Zinc					15000	14000	95	0			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133883 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV014				ID: CCV089				ID: CCV090		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	97	1	400000	410000	102	0	400000	100	0
Antimony	1000	910	91	0	5000	4900	98	0	4800	97	0
Arsenic	1000	1000	104	0	5000	5000	100	1	4900	99	0
Barium	520	520	100	1	10000	10000	104	0	10000	101	0
Beryllium	510	520	102	1	500	510	103	0	500	101	1
Cadmium	510	480	93	0	2500	2500	102	0	2500	100	0
Calcium	10000	10000	102	2	400000	410000	104	0	410000	101	0
Chromium	520	550	106	0	15500	16000	102	0	16000	101	0
Cobalt	520	500	96	0	2500	2600	102	1	2500	101	0
Copper	510	500	98	2	15000	15000	102	0	15000	100	0
Iron	10000	11000	106	1	400000	420000	104	0	410000	102	0
Lead	1000	970	97	0	25000	25000	101	0	25000	100	0
Magnesium	6000	6000	101	1	400000	420000	105	0	410000	102	0
Manganese	520	530	101	1	15000	15000	102	0	15000	100	0
Nickel	530	510	96	0	2500	2600	103	0	2500	101	0
Potassium	9900	9400	95	4	150000	160000	104	0	150000	102	0
Selenium	1000	930	93	0	5000	4900	97	0	4800	95	1
Silver	250	250	99	0	1250	1300	102	0	1300	100	0
Sodium	10000	10000	101	2	400000	410000	102	0	400000	100	0
Thallium	1000	980	98	0	5000	5000	100	0	4900	98	0
Vanadium	500	510	101	1	2500	2600	102	0	2500	101	0
Zinc	1000	1100	110	1	15000	15000	99	1	15000	99	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133883 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV091				ID: CCV092		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	1	400000	100	1
Antimony					5000	4800	96	1	4800	97	0
Arsenic					5000	4800	97	0	4900	98	0
Barium					10000	10000	100	1	10000	100	1
Beryllium					500	500	100	2	500	100	1
Cadmium					2500	2500	99	0	2500	100	0
Calcium					400000	400000	100	1	410000	102	1
Chromium					15500	15000	100	0	16000	101	1
Cobalt					2500	2500	101	0	2500	101	0
Copper					15000	15000	99	1	15000	100	1
Iron					400000	400000	101	1	410000	102	1
Lead					25000	25000	99	0	25000	99	0
Magnesium					400000	400000	99	1	400000	101	1
Manganese					15000	15000	97	1	15000	98	1
Nickel					2500	2500	100	0	2500	101	0
Potassium					150000	150000	100	1	150000	102	1
Selenium					5000	4700	94	0	4800	95	0
Silver					1250	1200	99	0	1300	100	0
Sodium					400000	390000	99	1	400000	99	1
Thallium					5000	4900	98	0	4900	98	0
Vanadium					2500	2500	100	1	2500	101	1
Zinc					15000	14000	95	1	14000	93	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133883 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV093				ID: CCV094		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	0	400000	99	0
Antimony					5000	4800	96	0	4800	97	0
Arsenic					5000	4900	98	0	5000	100	0
Barium					10000	10000	100	0	10000	100	0
Beryllium					500	500	100	1	500	100	0
Cadmium					2500	2500	99	0	2500	100	0
Calcium					400000	400000	101	0	400000	100	0
Chromium					15500	15000	100	0	15000	99	1
Cobalt					2500	2500	100	0	2500	101	0
Copper					15000	15000	100	0	15000	99	0
Iron					400000	400000	100	0	400000	100	0
Lead					25000	25000	99	0	25000	99	0
Magnesium					400000	400000	100	0	400000	99	0
Manganese					15000	15000	97	0	14000	96	0
Nickel					2500	2500	100	0	2500	101	0
Potassium					150000	150000	101	0	150000	101	0
Selenium					5000	4700	94	0	4700	95	0
Silver					1250	1200	99	0	1200	98	1
Sodium					400000	400000	99	0	390000	98	0
Thallium					5000	4900	97	0	4900	98	0
Vanadium					2500	2500	101	1	2500	100	0
Zinc					15000	14000	92	1	14000	91	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133949 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV007				ID: CCV008		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	95	0	400000	380000	94	0	380000	94	0
Antimony	1000	920	92	0	5000	4600	92	0	4600	92	0
Arsenic	1000	1100	109	0	5000	4700	95	0	4800	97	0
Barium	520	530	101	0	10000	9700	97	0	9700	97	0
Beryllium	510	520	103	0	500	480	95	0	480	95	0
Cadmium	510	500	97	0	2500	2400	94	0	2400	95	0
Calcium	10000	10000	102	1	400000	390000	98	0	390000	97	0
Chromium	520	560	107	1	15500	15000	95	0	15000	95	0
Cobalt	520	500	96	0	2500	2300	94	0	2400	94	0
Copper	510	510	100	0	15000	15000	98	0	15000	97	0
Iron	10000	11000	105	1	400000	390000	98	0	390000	97	0
Lead	1000	970	97	0	25000	23000	94	0	24000	95	0
Magnesium	6000	6200	104	0	400000	380000	95	0	380000	94	0
Manganese	520	540	104	0	15000	15000	98	0	14000	96	0
Nickel	530	510	96	0	2500	2400	95	0	2400	96	0
Potassium	9900	9500	96	2	150000	140000	96	0	140000	96	0
Selenium	1000	940	94	0	5000	4500	90	0	4600	92	0
Silver	250	260	103	0	1250	1200	98	0	1200	97	0
Sodium	10000	11000	106	3	400000	380000	95	0	380000	95	0
Thallium	1000	1000	104	1	5000	4800	97	0	4800	97	0
Vanadium	500	510	103	1	2500	2400	97	1	2400	97	0
Zinc	1000	1100	110	0	15000	14000	93	0	14000	93	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133949 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV009				ID: CCV010		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	380000	94	0	380000	96	0
Antimony					5000	4700	94	1	4600	92	0
Arsenic					5000	4800	97	0	4800	96	0
Barium					10000	9700	97	0	9700	97	0
Beryllium					500	480	96	0	480	96	1
Cadmium					2500	2400	95	0	2400	95	0
Calcium					400000	390000	97	1	390000	99	1
Chromium					15500	15000	94	0	15000	94	1
Cobalt					2500	2400	94	0	2300	93	0
Copper					15000	15000	97	0	15000	99	0
Iron					400000	390000	97	0	390000	99	0
Lead					25000	24000	95	0	24000	95	0
Magnesium					400000	380000	94	0	380000	95	1
Manganese					15000	14000	96	0	15000	98	0
Nickel					2500	2400	96	0	2400	96	0
Potassium					150000	140000	96	0	140000	95	0
Selenium					5000	4500	90	0	4500	90	0
Silver					1250	1200	97	0	1200	98	1
Sodium					400000	380000	94	1	380000	96	1
Thallium					5000	4800	97	1	4900	99	0
Vanadium					2500	2400	97	1	2500	98	0
Zinc					15000	14000	91	0	14000	91	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133949 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV011				ID: CCV012		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	380000	95	0	380000	95	0
Antimony					5000	4700	93	0	4700	94	2
Arsenic					5000	4900	98	0	4900	98	1
Barium					10000	9800	98	0	9700	97	0
Beryllium					500	480	95	0	480	96	1
Cadmium					2500	2400	96	0	2400	96	2
Calcium					400000	400000	100	0	400000	100	0
Chromium					15500	15000	95	0	15000	94	0
Cobalt					2500	2300	93	0	2300	93	2
Copper					15000	15000	99	0	15000	99	0
Iron					400000	400000	99	0	390000	99	0
Lead					25000	24000	96	0	24000	96	2
Magnesium					400000	380000	94	0	370000	94	0
Manganese					15000	15000	99	0	15000	98	0
Nickel					2500	2400	96	0	2400	96	2
Potassium					150000	140000	95	0	140000	94	0
Selenium					5000	4600	92	0	4600	92	2
Silver					1250	1200	100	0	1200	99	0
Sodium					400000	380000	96	0	380000	95	1
Thallium					5000	5100	102	0	5100	102	2
Vanadium					2500	2500	98	1	2500	99	1
Zinc					15000	14000	93	0	14000	91	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHHF2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133949 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV013				ID: CCV014		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	380000	95	3	390000	97	0
Antimony					5000	4800	96	1	4800	97	1
Arsenic					5000	5100	101	0	5000	100	0
Barium					10000	9900	99	3	10000	100	1
Beryllium					500	480	97	3	490	98	1
Cadmium					2500	2400	98	1	2400	98	0
Calcium					400000	400000	100	3	410000	101	0
Chromium					15500	15000	98	2	15000	97	0
Cobalt					2500	2400	95	1	2400	95	1
Copper					15000	15000	100	3	15000	101	1
Iron					400000	390000	98	3	390000	99	1
Lead					25000	25000	99	1	25000	98	1
Magnesium					400000	380000	94	3	380000	96	1
Manganese					15000	15000	100	3	15000	100	1
Nickel					2500	2500	98	1	2500	99	0
Potassium					150000	140000	95	3	140000	97	1
Selenium					5000	4700	95	1	4700	95	1
Silver					1250	1300	104	1	1300	103	0
Sodium					400000	380000	94	3	380000	96	1
Thallium					5000	5300	105	1	5200	105	0
Vanadium					2500	2500	100	3	2500	101	1
Zinc					15000	14000	93	2	14000	92	0