

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51882MA No.: _____ SDG No.: MX1007Instrument ID: P5Start Date : 12/02/2024Analytical Method: ICP-AESRun Batch : LB133698

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999908	0.079563	-0.002282	WLR Blank	Inverse Conc
Antimony	0.997678	0.156333	-0.000028	WLR Blank	Inverse Conc
Arsenic	0.998969	0.027796	-0.000161	WLR Blank	Inverse Conc
Barium	0.999940	3.543739	-0.002067	WLR Blank	Inverse Conc
Beryllium	0.999861	3.389273	-0.000953	WLR Blank	Inverse Conc
Boron	0.999064	0.212099	0.002406	WLR Blank	Inverse Conc
Cadmium	0.999986	1.807009	0.000353	WLR Blank	Inverse Conc
Calcium	0.999897	0.069559	0.004286	WLR Blank	Inverse Conc
Chromium	0.999752	0.116066	-0.000202	WLR Blank	Inverse Conc
Cobalt	0.999963	0.543535	-0.000318	WLR Blank	Inverse Conc
Copper	0.999974	0.202778	0.001848	WLR Blank	Inverse Conc
Iron	0.999948	0.018929	0.000261	WLR Blank	Inverse Conc
Lead	0.999970	0.132662	-0.000205	WLR Blank	Inverse Conc
Lithium	0.999645	1.357791	0.067757	WLR Blank	Inverse Conc
Magnesium	0.999839	0.007394	-0.000095	WLR Blank	Inverse Conc
Manganese	0.999905	0.498854	-0.000165	WLR Blank	Inverse Conc
Molybdenum	0.999065	0.715331	-0.000179	WLR Blank	Inverse Conc
Nickel	0.999952	0.534681	0.000049	WLR Blank	Inverse Conc
Phosphorus	0.999290	0.043186	-0.001290	WLR Blank	Inverse Conc
Potassium	0.999688	0.018597	0.077967	WLR Blank	Inverse Conc
Selenium	0.998329	0.055397	-0.000014	WLR Blank	Inverse Conc
Silicon	0.998483	0.786594	0.004970	WLR Blank	Inverse Conc
Silver	0.999754	0.328605	0.020745	WLR Blank	Inverse Conc
Sodium	0.999929	0.005421	-0.003534	WLR Blank	Inverse Conc
Strontium	0.999910	8.380198	0.000828	WLR Blank	Inverse Conc
Sulfur	0.998963	0.025168	0.000222	WLR Blank	Inverse Conc
Thallium	0.999854	0.055843	-0.000374	WLR Blank	Inverse Conc
Tin	0.999936	0.090218	0.000198	WLR Blank	Inverse Conc
Vanadium	0.999839	0.085442	-0.000726	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51882 MA No.: SDG No.: MX1007
Instrument ID: P5 Start Date : 12/02/2024
Analytical Method: ICP-AES Run Batch : LB133698

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.998429	2.086653	0.002237	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51882 MA No.: SDG No.: MX1007
Instrument ID: P5 Start Date : 12/02/2024
Analytical Method: ICP-AES Run Batch : LB133698

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------