

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51882 MA No. :                      SDG No.: MX1007  
 Initial Calibration Verification Source : WP110845  
 Continuing Calibration Verification Source : WP110844  
 Run Batch: LB133577 Analytical Method: Spectrophotometry  
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Cyanide	99.0	96	97		250	250	99		250	100	