

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51882 MA No.:                      SDG No.: MX1007  
Instrument ID: CV1 Start Date: 12/03/2024  
Analytical Method: CVAA Run Batch: LB133714  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D | True | Found | %D |
|---------|------|-------|----|------|-------|----|------|-------|----|
| Mercury | 0    | 0.0   | 0  | 0.2  | 0.19  | -5 | 2.5  | 2.6   | 3  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51882 MA No.: \_\_\_\_\_ SDG No.: MX1007  
Instrument ID: CV1 Start Date: 12/03/2024  
Analytical Method: CVAA Run Batch: LB133714  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D | True | Found | %D |
|---------|------|-------|----|------|-------|----|------|-------|----|
| Mercury | 5    | 4.9   | -2 | 7.5  | 7.6   | 2  | 10   | 9.9   | -1 |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51882 MA No.: \_\_\_\_\_ SDG No.: MX1007  
Instrument ID: CV1 Start Date: 12/04/2024  
Analytical Method: CVAA Run Batch: LB133725  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D | True | Found | %D |
|---------|------|-------|----|------|-------|----|------|-------|----|
| Mercury | 0    | 0.024 | 0  | 0.2  | 0.21  | 4  | 2.5  | 2.5   | 2  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51882 MA No.: \_\_\_\_\_ SDG No.: MX1007  
Instrument ID: CV1 Start Date: 12/04/2024  
Analytical Method: CVAA Run Batch: LB133725  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D | True | Found | %D |
|---------|------|-------|----|------|-------|----|------|-------|----|
| Mercury | 5    | 4.9   | -2 | 7.5  | 7.5   | -1 | 10   | 10    | 1  |