

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51882 MA No.: SDG No.: MX1007
Instrument ID: CV1 Start Date : 12/03/2024
Analytical Method: CVAA Run Batch : LB133714

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Mercury	0.999732	0.000251	-0.080939	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51882 MA No.: SDG No.: MX1007
Instrument ID: CV1 Start Date : 12/04/2024
Analytical Method: CVAA Run Batch : LB133725

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Mercury	0.999878	0.000227	-0.019797	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51882 MA No.: SDG No.: MX1007
Instrument ID: CV1 Start Date : 12/04/2024
Analytical Method: CVAA Run Batch : LB133725

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------