

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

Continuing Calib Date: 11/20/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 12:20 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	5.92	5.92	5.82	6.02	0.00
Aroclor-1242-2	(2)	5.95	5.95	5.85	6.05	0.00
Aroclor-1242-3	(3)	6.01	6.01	5.91	6.11	0.00
Aroclor-1242-4	(4)	6.11	6.11	6.01	6.21	0.00
Aroclor-1242-5	(5)	6.84	6.84	6.74	6.94	0.00
Tetrachloro-m-xylene		4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl		10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

Continuing Calib Date: 11/20/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 12:20 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1242-2	(2)	5.18	5.18	5.08	5.28	0.00
Aroclor-1242-3	(3)	5.36	5.36	5.26	5.46	0.00
Aroclor-1242-4	(4)	5.44	5.44	5.34	5.54	0.00
Aroclor-1242-5	(5)	5.97	5.97	5.87	6.07	0.00
Tetrachloro-m-xylene		4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl		9.21	9.21	9.11	9.31	0.00