

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

Continuing Calib Date: 11/20/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 12:36 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.93	5.92	5.82	6.02	-0.01
Aroclor-1248-2 (2)	6.20	6.19	6.09	6.29	-0.01
Aroclor-1248-3 (3)	6.41	6.40	6.30	6.50	-0.01
Aroclor-1248-4 (4)	6.81	6.80	6.70	6.90	0.00
Aroclor-1248-5 (5)	6.85	6.84	6.74	6.94	0.00
Tetrachloro-m-xylene	4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl	10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

Continuing Calib Date: 11/20/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 12:36 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1248-2 (2)	5.40	5.40	5.30	5.50	0.00
Aroclor-1248-3 (3)	5.44	5.44	5.34	5.54	0.00
Aroclor-1248-4 (4)	5.62	5.62	5.52	5.72	0.01
Aroclor-1248-5 (5)	6.01	6.01	5.91	6.11	0.00
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.21	9.21	9.11	9.31	0.00