

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PP068617.D Time Analyzed: 12:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.925	5.821	6.021	471.490	500.000	-5.7
Aroclor-1016-2	5.947	5.844	6.044	473.640	500.000	-5.3
Aroclor-1016-3	6.011	5.908	6.108	483.080	500.000	-3.4
Aroclor-1016-4	6.109	6.005	6.205	493.960	500.000	-1.2
Aroclor-1016-5	6.403	6.299	6.499	473.390	500.000	-5.3
Aroclor-1260-1	7.527	7.423	7.623	451.860	500.000	-9.6
Aroclor-1260-2	7.780	7.676	7.876	446.670	500.000	-10.7
Aroclor-1260-3	8.143	8.038	8.238	434.600	500.000	-13.1
Aroclor-1260-4	8.380	8.276	8.476	429.660	500.000	-14.1
Aroclor-1260-5	8.718	8.613	8.813	430.310	500.000	-13.9
Decachlorobiphenyl	10.679	10.572	10.772	44.530	50.000	-10.9
Tetrachloro-m-xylene	4.760	4.657	4.857	48.550	50.000	-2.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PP068617.D Time Analyzed: 12:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.156	5.058	5.258	491.600	500.000	-1.7
Aroclor-1016-2	5.176	5.079	5.279	479.260	500.000	-4.1
Aroclor-1016-3	5.357	5.259	5.459	468.720	500.000	-6.3
Aroclor-1016-4	5.396	5.298	5.498	476.580	500.000	-4.7
Aroclor-1016-5	5.615	5.517	5.717	472.210	500.000	-5.6
Aroclor-1260-1	6.658	6.560	6.760	453.610	500.000	-9.3
Aroclor-1260-2	6.845	6.747	6.947	443.010	500.000	-11.4
Aroclor-1260-3	7.002	6.905	7.105	457.630	500.000	-8.5
Aroclor-1260-4	7.476	7.378	7.578	461.930	500.000	-7.6
Aroclor-1260-5	7.715	7.617	7.817	460.650	500.000	-7.9
Decachlorobiphenyl	9.207	9.110	9.310	45.410	50.000	-9.2
Tetrachloro-m-xylene	4.051	3.951	4.151	49.260	50.000	-1.5