

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PP068632.D Time Analyzed: 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.924	5.821	6.021	479.130	500.000	-4.2
Aroclor-1016-2	5.946	5.844	6.044	483.850	500.000	-3.2
Aroclor-1016-3	6.009	5.908	6.108	498.710	500.000	-0.3
Aroclor-1016-4	6.108	6.005	6.205	509.210	500.000	1.8
Aroclor-1016-5	6.402	6.299	6.499	475.890	500.000	-4.8
Aroclor-1260-1	7.526	7.423	7.623	456.980	500.000	-8.6
Aroclor-1260-2	7.779	7.676	7.876	454.010	500.000	-9.2
Aroclor-1260-3	8.141	8.038	8.238	445.410	500.000	-10.9
Aroclor-1260-4	8.379	8.276	8.476	435.660	500.000	-12.9
Aroclor-1260-5	8.717	8.613	8.813	434.470	500.000	-13.1
Decachlorobiphenyl	10.677	10.572	10.772	45.380	50.000	-9.2
Tetrachloro-m-xylene	4.759	4.657	4.857	48.700	50.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PP068632.D Time Analyzed: 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.058	5.258	484.670	500.000	-3.1
Aroclor-1016-2	5.177	5.079	5.279	474.220	500.000	-5.2
Aroclor-1016-3	5.358	5.259	5.459	465.110	500.000	-7.0
Aroclor-1016-4	5.397	5.298	5.498	473.950	500.000	-5.2
Aroclor-1016-5	5.616	5.517	5.717	472.100	500.000	-5.6
Aroclor-1260-1	6.659	6.560	6.760	458.030	500.000	-8.4
Aroclor-1260-2	6.845	6.747	6.947	451.460	500.000	-9.7
Aroclor-1260-3	7.003	6.905	7.105	459.830	500.000	-8.0
Aroclor-1260-4	7.477	7.378	7.578	473.280	500.000	-5.3
Aroclor-1260-5	7.716	7.617	7.817	460.560	500.000	-7.9
Decachlorobiphenyl	9.208	9.110	9.310	46.310	50.000	-7.4
Tetrachloro-m-xylene	4.051	3.951	4.151	49.260	50.000	-1.5