

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/21/2024

Lab Sample No.: AR1660CCC500 Data File : PP068647.D Time Analyzed: 01:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.928	5.821	6.021	486.810	500.000	-2.6
Aroclor-1016-2	5.950	5.844	6.044	500.230	500.000	0.0
Aroclor-1016-3	6.013	5.908	6.108	522.410	500.000	4.5
Aroclor-1016-4	6.112	6.005	6.205	524.880	500.000	5.0
Aroclor-1016-5	6.406	6.299	6.499	471.670	500.000	-5.7
Aroclor-1260-1	7.531	7.423	7.623	450.950	500.000	-9.8
Aroclor-1260-2	7.784	7.676	7.876	445.870	500.000	-10.8
Aroclor-1260-3	8.145	8.038	8.238	435.910	500.000	-12.8
Aroclor-1260-4	8.385	8.276	8.476	428.030	500.000	-14.4
Aroclor-1260-5	8.721	8.613	8.813	427.630	500.000	-14.5
Decachlorobiphenyl	10.683	10.572	10.772	44.670	50.000	-10.7
Tetrachloro-m-xylene	4.763	4.657	4.857	49.950	50.000	-0.1

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4919 SAS No.: P4919 SDG NO.: P4919

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/21/2024

Lab Sample No.: AR1660CCC500 Data File : PP068647.D Time Analyzed: 01:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.157	5.058	5.258	470.050	500.000	-6.0
Aroclor-1016-2	5.177	5.079	5.279	486.040	500.000	-2.8
Aroclor-1016-3	5.358	5.259	5.459	490.540	500.000	-1.9
Aroclor-1016-4	5.398	5.298	5.498	498.600	500.000	-0.3
Aroclor-1016-5	5.616	5.517	5.717	478.020	500.000	-4.4
Aroclor-1260-1	6.659	6.560	6.760	465.510	500.000	-6.9
Aroclor-1260-2	6.846	6.747	6.947	455.110	500.000	-9.0
Aroclor-1260-3	7.004	6.905	7.105	465.310	500.000	-6.9
Aroclor-1260-4	7.478	7.378	7.578	472.800	500.000	-5.4
Aroclor-1260-5	7.717	7.617	7.817	459.670	500.000	-8.1
Decachlorobiphenyl	9.210	9.110	9.310	45.990	50.000	-8.0
Tetrachloro-m-xylene	4.052	3.951	4.151	49.470	50.000	-1.1