

Surrogate Summary

SDG No.: **P4919**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP068360.D	PIBLK-PP068360.D	Tetrachloro-m-xylene	1	20	21.0	105		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
I.BLK-PP068621.D	PIBLK-PP068621.D	Tetrachloro-m-xylene	1	20	21.4	107		60	140
		Decachlorobiphenyl	1	20	22.4	112		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	22.5	113		60	140
PB165124BL	PB165124BL	Tetrachloro-m-xylene	1	20	19.6	98		32	144
		Decachlorobiphenyl	1	20	20.8	104		32	175
		Tetrachloro-m-xylene	2	20	19.4	97		32	144
		Decachlorobiphenyl	2	20	20.9	104		32	175
PB165124BS	PB165124BS	Tetrachloro-m-xylene	1	20	20.7	104		32	144
		Decachlorobiphenyl	1	20	21.5	108		32	175
		Tetrachloro-m-xylene	2	20	20.6	103		32	144
		Decachlorobiphenyl	2	20	21.9	110		32	175
P4919-01	ROOF-1-OFF-WHITE	Tetrachloro-m-xylene	1	20	16.9	84		32	144
		Decachlorobiphenyl	1	20	20.5	103		32	175
		Tetrachloro-m-xylene	2	20	20.8	104		32	144
		Decachlorobiphenyl	2	20	21.7	108		32	175
P4919-03	ROOF-11	Tetrachloro-m-xylene	1	20	19.5	97		32	144
		Decachlorobiphenyl	1	20	19.3	96		32	175
		Tetrachloro-m-xylene	2	20	19.6	98		32	144
		Decachlorobiphenyl	2	20	20.4	102		32	175
I.BLK-PP068636.D	PIBLK-PP068636.D	Tetrachloro-m-xylene	1	20	21.4	107		60	140
		Decachlorobiphenyl	1	20	22.8	114		60	140
		Tetrachloro-m-xylene	2	20	21.9	109		60	140
		Decachlorobiphenyl	2	20	23.2	116		60	140
P4919-02	ROOF-1-BLACK	Tetrachloro-m-xylene	1	20	19.6	98		32	144
		Decachlorobiphenyl	1	20	26.0	130		32	175
		Tetrachloro-m-xylene	2	20	12.9	65		32	144
		Decachlorobiphenyl	2	20	19.0	95		32	175
I.BLK-PP068651.D	PIBLK-PP068651.D	Tetrachloro-m-xylene	1	20	21.6	108		60	140
		Decachlorobiphenyl	1	20	22.4	112		60	140
		Tetrachloro-m-xylene	2	20	21.9	109		60	140
		Decachlorobiphenyl	2	20	23.0	115		60	140