



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4921

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133526

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	11/20/2024
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	11/20/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/20/2024

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4921

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133527

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.4	81	102	78-84	11/20/2024

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4921

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133548

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.096	0.099	97	85-115	11/21/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/21/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	11/21/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	11/21/2024

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4921

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133553

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Cyanide		mg/L	0.096	0.099	97	90-110	11/21/2024
Sample ID:	CCV1						
Cyanide		mg/L	0.25	0.25	100	90-110	11/21/2024
Sample ID:	CCV2						
Cyanide		mg/L	0.25	0.25	100	90-110	11/21/2024