

### Laboratory Control Sample Summary

**Client:** AECOM

**SDG No.:** P4921

**Project:** Meeker Ave Plumes Superfund Site RI FS

**Run No.:** LB133553

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB165170BS |            |        |                 |            |                 |                     |               |
| Cyanide   | mg/L       | 0.1        | 0.095  |                 | 95         | 1               | 85-115              | 11/21/2024    |