


LB133570

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	-27	0.000	-27	0				1
0.2	0.200	0.223	0.023	840	0.0 %	840					2
2.5	2.500	2.547	0.047	9528	0.0 %	9528					2
5.0	5.000	5.087	0.087	19027	0.0 %	19027					2
7.5	7.500	7.183	-0.317	26865	0.0 %	26865					2
10.0	10.000	10.168	0.168	38028	0.0 %	38028					2

LB133570 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Nov 2024 08:28:38

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	-27	-	0.000	7470A	PPB	22 Nov 2024 08:29:27	S	Std
0.2 - 1	50.2	840	-	0.200	7470A	PPB	22 Nov 2024 08:31:44	S	Std
2.5 - 1	52.5	9528	-	2.500	7470A	PPB	22 Nov 2024 08:34:01	S	Std
5.0 - 1	55	19027	-	5.000	7470A	PPB	22 Nov 2024 08:36:18	S	Std
7.5 - 1	59.5	26865	-	7.500	7470A	PPB	22 Nov 2024 08:38:35	S	Std
10.0 - 1	510	38028	-	10.000	7470A	PPB	22 Nov 2024 08:47:43	S	Std
ICV97 - 1	ICV97	13647	3.6483	-	7470A	PPB	22 Nov 2024 08:52:58	U	SMPL
ICB97 - 1	ICB97	-236	-0.0643	-	7470A	PPB	22 Nov 2024 08:55:12	U	SMPL
CCV31 - 1	CCV31	19532	5.2220	-	7470A	PPB	22 Nov 2024 08:57:32	U	SMPL
CCB31 - 1	CCB31	-324	-0.0878	-	7470A	PPB	22 Nov 2024 08:59:47	U	SMPL
CRA - 1	CRA	735	0.1954	-	7470A	PPB	22 Nov 2024 09:02:04	U	SMPL
HighStd - 1	HighStd	36675	9.8064	-	7470A	PPB	22 Nov 2024 09:04:19	U	SMPL
ChkStd - 1	ChkStd	25582	6.8399	-	7470A	PPB	22 Nov 2024 09:09:24	U	SMPL
PB165143BL - 1	PBW	-86	-0.0242	-	7470A	PPB	22 Nov 2024 09:14:14	U	SMPL
PB165143BS - 1	LCSW	16144	4.3160	-	7470A	PPB	22 Nov 2024 09:16:29	U	SMPL
P4921-01 - 1	WC-11-A-202411	-253	-0.0688	-	7470A	PPB	22 Nov 2024 09:18:45	U	SMPL
P4921-01DUP - 1	WC-11-A-202411DUP	-99	-0.0276	-	7470A	PPB	22 Nov 2024 09:21:03	U	SMPL
P4921-01MS - 1	WC-11-A-202411MS	14084	3.7651	-	7470A	PPB	22 Nov 2024 09:23:20	U	SMPL
P4921-01MSD - 1	WC-11-A-202411MSD	14539	3.8868	-	7470A	PPB	22 Nov 2024 09:25:37	U	SMPL
PB165179BL - 1	PBW	-313	-0.0849	-	7470A	PPB	22 Nov 2024 09:27:54	U	SMPL
CCV32 - 1	CCV32	18814	5.0300	-	7470A	PPB	22 Nov 2024 09:30:11	U	SMPL
CCB32 - 1	CCB32	-328	-0.0889	-	7470A	PPB	22 Nov 2024 09:32:26	U	SMPL
PB165179BS - 1	LCSW	16456	4.3994	-	7470A	PPB	22 Nov 2024 09:34:44	U	SMPL
P4916-04 - 1	TP-1-WC	-273	-0.0742	-	7470A	PPB	22 Nov 2024 09:36:59	U	SMPL
P4916-08 - 1	TP-2-WC	-92	-0.0258	-	7470A	PPB	22 Nov 2024 09:39:16	U	SMPL
P4916-12 - 1	TP-3-WC	-68	-0.0194	-	7470A	PPB	22 Nov 2024 09:41:31	U	SMPL
P4923-02 - 1	COMP-1	-74	-0.0210	-	7470A	PPB	22 Nov 2024 09:43:46	U	SMPL
P4923-03 - 1	COMP-2	-34	-0.0103	-	7470A	PPB	22 Nov 2024 09:46:03	U	SMPL
P4923-04 - 1	COMP-3	-40	-0.0119	-	7470A	PPB	22 Nov 2024 09:48:19	U	SMPL
P4923-05 - 1	COMP-4	-47	-0.0137	-	7470A	PPB	22 Nov 2024 09:50:35	U	SMPL
P4923-07 - 1	72-11991	-102	-0.0285	-	7470A	PPB	22 Nov 2024 09:52:52	U	SMPL
P4924-04 - 1	MH-4	-61	-0.0175	-	7470A	PPB	22 Nov 2024 09:55:07	U	SMPL
CCV33 - 1	CCV33	18264	4.8829	-	7470A	PPB	22 Nov 2024 09:57:22	U	SMPL
CCB33 - 1	CCB33	-265	-0.0720	-	7470A	PPB	22 Nov 2024 09:59:37	U	SMPL
P4925-04 - 1	MH-741	-73	-0.0207	-	7470A	PPB	22 Nov 2024 10:01:55	U	SMPL
P4925-08 - 1	MH-741	-72	-0.0204	-	7470A	PPB	22 Nov 2024 10:04:10	U	SMPL
P4929-02 - 1	ARS520	-42	-0.0124	-	7470A	PPB	22 Nov 2024 10:06:26	U	SMPL
P4929-02DUP - 1	ARS520DUP	-58	-0.0167	-	7470A	PPB	22 Nov 2024 10:08:41	U	SMPL
P4929-02MS - 1	ARS520MS	14375	3.8430	-	7470A	PPB	22 Nov 2024 10:10:56	U	SMPL
P4929-02MSD - 1	ARS520MSD	15387	4.1136	-	7470A	PPB	22 Nov 2024 10:13:12	U	SMPL
PB165183BL - 1	PBW	-277	-0.0753	-	7470A	PPB	22 Nov 2024 10:15:30	U	SMPL
PB165183BS - 1	LCSW	16487	4.4077	-	7470A	PPB	22 Nov 2024 10:17:49	U	SMPL
P4892-04 - 1	WB-310-SW	-129	-0.0357	-	7470A	PPB	22 Nov 2024 10:20:06	U	SMPL
P4892-04DUP - 1	WB-310-SWDUP	-10	-0.0038	-	7470A	PPB	22 Nov 2024 10:22:23	U	SMPL
CCV34 - 1	CCV34	18398	4.9188	-	7470A	PPB	22 Nov 2024 10:27:24	U	SMPL
CCB34 - 1	CCB34	-159	-0.0437	-	7470A	PPB	22 Nov 2024 10:31:07	U	SMPL
P4892-04MS - 1	WB-310-SWMS	16268	4.3492	-	7470A	PPB	22 Nov 2024 10:33:22	U	SMPL
P4892-04MSD - 1	WB-310-SWMSD	15418	4.1219	-	7470A	PPB	22 Nov 2024 10:35:37	U	SMPL
P4927-01 - 1	111424-C	332	0.0876	-	7470A	PPB	22 Nov 2024 10:37:55	U	SMPL
PB165111TB - 1	PB165111TB	138	0.0357	-	7470A	PPB	22 Nov 2024 10:42:50	U	SMPL
PB165123TB - 1	PB165123TB	-23	-0.0073	-	7470A	PPB	22 Nov 2024 10:45:06	U	SMPL
CCV35 - 1	CCV35	20177	5.3945	-	7470A	PPB	22 Nov 2024 10:47:21	U	SMPL
CCB35 - 1	CCB35	-235	-0.0640	-	7470A	PPB	22 Nov 2024 10:52:20	U	SMPL
P4921-01LX5 - 1		-62	-0.0178	-	7470A	PPB	22 Nov 2024 10:54:36	U	SMPL
P4921-01A - 1		16106	4.3059	-	7470A	PPB	22 Nov 2024 10:56:53	U	SMPL
P4929-02LX5 - 1		-202	-0.0552	-	7470A	PPB	22 Nov 2024 10:59:08	U	SMPL
P4929-02A - 1		14842	3.9678	-	7470A	PPB	22 Nov 2024 11:01:25	U	SMPL
P4892-04LX5 - 1		-230	-0.0627	-	7470A	PPB	22 Nov 2024 11:03:41	U	SMPL
P4892-04A - 1		16074	4.2973	-	7470A	PPB	22 Nov 2024 11:05:58	U	SMPL
CCV36 - 1	CCV36	19861	5.3100	-	7470A	PPB	22 Nov 2024 11:17:06	U	SMPL
CCB36 - 1	CCB36	-335	-0.0908	-	7470A	PPB	22 Nov 2024 11:19:21	U	SMPL