



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: AECOM **SDG No.:** P4921
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4921 **SAS No.:** P4921
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	21.2	20.0	106	40 - 160	P	11/21/2024	12:45	LB133568
	Barium	96.2	100	96	40 - 160	P	11/21/2024	12:45	LB133568
	Cadmium	7.10	6.0	118	40 - 160	P	11/21/2024	12:45	LB133568
	Chromium	10.2	10.0	102	40 - 160	P	11/21/2024	12:45	LB133568
	Lead	11.6	12.0	97	40 - 160	P	11/21/2024	12:45	LB133568
	Selenium	20.0	20.0	100	40 - 160	P	11/21/2024	12:45	LB133568
	Silver	10.6	10.0	106	40 - 160	P	11/21/2024	12:45	LB133568
CRA	Mercury	0.20	0.2	98	40 - 160	CV	11/22/2024	09:02	LB133570