

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: AECOM **SDG No.:** P4921
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4921 **SAS No.:** P4921
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.85			-20	20	11/21/2024	12:49	LB133568
	Barium	2.20	6.0	37	-94	106	11/21/2024	12:49	LB133568
	Cadmium	6.13	1.0	613	-5	7	11/21/2024	12:49	LB133568
	Chromium	59.3	52.0	114	42	62	11/21/2024	12:49	LB133568
	Lead	7.66			-12	12	11/21/2024	12:49	LB133568
	Selenium	-7.87			-20	20	11/21/2024	12:49	LB133568
	Silver	-2.87			-10	10	11/21/2024	12:49	LB133568
ICSAB01	Arsenic	117	104	112	88.4	120	11/21/2024	12:58	LB133568
	Barium	511	537	95	437	637	11/21/2024	12:58	LB133568
	Cadmium	1030	972	106	826	1120	11/21/2024	12:58	LB133568
	Chromium	575	542	106	460	624	11/21/2024	12:58	LB133568
	Lead	59.0	49.0	120	37	61	11/21/2024	12:58	LB133568
	Selenium	43.0	46.0	94	26	66	11/21/2024	12:58	LB133568
	Silver	201	201	100	170	232	11/21/2024	12:58	LB133568