

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** P4921
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4921 **SAS No.:** P4921
Matrix: Water **Sample ID:** P4921-01 **Client ID:** WC-11-A-202411DUP
Percent Solids for Sample: NA **Duplicate ID** P4921-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	2.00	U	2.00	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** P4921
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4921 **SAS No.:** P4921
Matrix: Water **Sample ID:** P4921-01MS **Client ID:** WC-11-A-202411MSD
Percent Solids for Sample: NA **Duplicate ID** P4921-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	37.7		38.9		3		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** P4921
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4921 **SAS No.:** P4921
Matrix: Water **Sample ID:** P4929-02 **Client ID:** ARS520DUP
Percent Solids for Sample: NA **Duplicate ID** P4929-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	827		875		6		P
Cadmium	ug/L	20	5.50	J	4.68	J	16		P
Chromium	ug/L	20	10.8	J	10.8	J	1		P
Lead	ug/L	20	52.7	J	49.9	J	5		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** P4921
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4921 **SAS No.:** P4921
Matrix: Water **Sample ID:** P4929-02MS **Client ID:** ARS520MSD
Percent Solids for Sample: NA **Duplicate ID** P4929-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3870		4080		5		P
Barium	ug/L	20	1740		1850		6		P
Cadmium	ug/L	20	918		972		6		P
Chromium	ug/L	20	1940		2050		6		P
Lead	ug/L	20	4460		4710		5		P
Selenium	ug/L	20	9560		10000		4		P
Silver	ug/L	20	353		375		6		P