

CALIBRATION VERIFICATION SUMMARY

Contract: AEEO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

Continuing Calib Date: 11/20/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 19:38 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.53	5.53	5.43	5.63	0.00
Aroclor-1016-2	(2)	5.55	5.55	5.45	5.65	0.00
Aroclor-1016-3	(3)	5.61	5.61	5.51	5.71	0.00
Aroclor-1016-4	(4)	5.71	5.71	5.61	5.81	0.00
Aroclor-1016-5	(5)	6.00	6.00	5.90	6.10	0.00
Aroclor-1260-1	(1)	7.14	7.14	7.04	7.24	0.00
Aroclor-1260-2	(2)	7.40	7.40	7.30	7.50	0.01
Aroclor-1260-3	(3)	7.76	7.76	7.66	7.86	0.00
Aroclor-1260-4	(4)	7.98	7.98	7.88	8.08	0.00
Aroclor-1260-5	(5)	8.30	8.30	8.20	8.40	0.00
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.09	10.09	9.99	10.19	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

Continuing Calib Date: 11/20/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 19:38 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1016-2	(2)	4.74	4.74	4.64	4.84	0.00
Aroclor-1016-3	(3)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.17	5.07	5.27	0.00
Aroclor-1260-1	(1)	6.20	6.20	6.10	6.30	0.00
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.54	6.54	6.44	6.64	0.00
Aroclor-1260-4	(4)	7.01	7.01	6.91	7.11	0.00
Aroclor-1260-5	(5)	7.25	7.25	7.15	7.35	0.00
Tetrachloro-m-xylene		3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01