

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108149.D Time Analyzed: 19:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.526	5.425	5.625	528.630	500.000	5.7
Aroclor-1016-2	5.549	5.447	5.647	520.240	500.000	4.0
Aroclor-1016-3	5.611	5.510	5.710	513.970	500.000	2.8
Aroclor-1016-4	5.708	5.607	5.807	521.170	500.000	4.2
Aroclor-1016-5	6.004	5.904	6.104	514.630	500.000	2.9
Aroclor-1260-1	7.138	7.037	7.237	518.810	500.000	3.8
Aroclor-1260-2	7.395	7.295	7.495	545.340	500.000	9.1
Aroclor-1260-3	7.759	7.658	7.858	520.170	500.000	4.0
Aroclor-1260-4	7.984	7.884	8.084	517.420	500.000	3.5
Aroclor-1260-5	8.300	8.198	8.398	525.890	500.000	5.2
Decachlorobiphenyl	10.093	9.991	10.191	50.960	50.000	1.9
Tetrachloro-m-xylene	4.371	4.270	4.470	53.830	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108149.D Time Analyzed: 19:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.620	4.820	547.380	500.000	9.5
Aroclor-1016-2	4.738	4.639	4.839	545.840	500.000	9.2
Aroclor-1016-3	4.914	4.815	5.015	540.160	500.000	8.0
Aroclor-1016-4	4.956	4.856	5.056	518.270	500.000	3.7
Aroclor-1016-5	5.168	5.070	5.270	554.520	500.000	10.9
Aroclor-1260-1	6.200	6.100	6.300	547.210	500.000	9.4
Aroclor-1260-2	6.387	6.288	6.488	534.890	500.000	7.0
Aroclor-1260-3	6.541	6.441	6.641	524.600	500.000	4.9
Aroclor-1260-4	7.012	6.912	7.112	520.030	500.000	4.0
Aroclor-1260-5	7.253	7.154	7.354	511.160	500.000	2.2
Decachlorobiphenyl	8.633	8.535	8.735	47.130	50.000	-5.7
Tetrachloro-m-xylene	3.638	3.539	3.739	55.980	50.000	12.0