

### CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108159.D Time Analyzed: 23:06

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.526  | 5.425     | 5.625  | 545.410            | 500.000           | 9.1  |
| Aroclor-1016-2       | 5.548  | 5.447     | 5.647  | 539.240            | 500.000           | 7.8  |
| Aroclor-1016-3       | 5.611  | 5.510     | 5.710  | 531.900            | 500.000           | 6.4  |
| Aroclor-1016-4       | 5.708  | 5.607     | 5.807  | 546.420            | 500.000           | 9.3  |
| Aroclor-1016-5       | 6.005  | 5.904     | 6.104  | 552.530            | 500.000           | 10.5 |
| Aroclor-1260-1       | 7.138  | 7.037     | 7.237  | 546.400            | 500.000           | 9.3  |
| Aroclor-1260-2       | 7.397  | 7.295     | 7.495  | 577.560            | 500.000           | 15.5 |
| Aroclor-1260-3       | 7.760  | 7.658     | 7.858  | 556.610            | 500.000           | 11.3 |
| Aroclor-1260-4       | 7.985  | 7.884     | 8.084  | 559.850            | 500.000           | 12.0 |
| Aroclor-1260-5       | 8.301  | 8.198     | 8.398  | 566.220            | 500.000           | 13.2 |
| Decachlorobiphenyl   | 10.097 | 9.991     | 10.191 | 53.910             | 50.000            | 7.8  |
| Tetrachloro-m-xylene | 4.371  | 4.270     | 4.470  | 56.130             | 50.000            | 12.3 |

### CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108159.D Time Analyzed: 23:06

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.719 | 4.620                | 4.820 | 547.280            | 500.000           | 9.5  |
| Aroclor-1016-2       | 4.738 | 4.639                | 4.839 | 549.370            | 500.000           | 9.9  |
| Aroclor-1016-3       | 4.914 | 4.815                | 5.015 | 546.590            | 500.000           | 9.3  |
| Aroclor-1016-4       | 4.956 | 4.856                | 5.056 | 532.480            | 500.000           | 6.5  |
| Aroclor-1016-5       | 5.169 | 5.070                | 5.270 | 557.300            | 500.000           | 11.5 |
| Aroclor-1260-1       | 6.200 | 6.100                | 6.300 | 549.930            | 500.000           | 10.0 |
| Aroclor-1260-2       | 6.388 | 6.288                | 6.488 | 546.610            | 500.000           | 9.3  |
| Aroclor-1260-3       | 6.541 | 6.441                | 6.641 | 538.450            | 500.000           | 7.7  |
| Aroclor-1260-4       | 7.013 | 6.912                | 7.112 | 537.050            | 500.000           | 7.4  |
| Aroclor-1260-5       | 7.254 | 7.154                | 7.354 | 536.090            | 500.000           | 7.2  |
| Decachlorobiphenyl   | 8.636 | 8.535                | 8.735 | 50.810             | 50.000            | 1.6  |
| Tetrachloro-m-xylene | 3.638 | 3.539                | 3.739 | 56.020             | 50.000            | 12.0 |