



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4921

Client: AECOM

Analytical Method: **8082A** **Datafile :** PO108156.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165133BSD	AR1016	5	4.90	ug/L	98	0			61	112	20
	AR1260	5	4.90	ug/L	98	0			66	113	20