

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB165164BS

Contract: AECO02

Lab Code: CHEM

Case No.: P4921

SAS No.: P4921

SDG NO.: P4921

Lab Sample ID: PB165164BS

Date(s) Analyzed: 11/21/2024

11/21/2024

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR2

ID: 0.32 (mm)

GC Column: (2): ZB-MR1

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.50	7.45	7.55	0.43	14
	2	6.61	6.56	6.66	0.50	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	0.48	4.1
	2	3.61	3.56	3.66	0.50	
Heptachlor	1	4.91	4.86	4.96	0.49	6.3
	2	3.95	3.90	4.00	0.52	
Heptachlor epoxide	1	5.68	5.63	5.73	0.47	11.6
	2	4.73	4.68	4.78	0.53	
Endrin	1	6.57	6.52	6.62	0.42	16.5
	2	5.64	5.59	5.69	0.49	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WB-310-BOTMS

Contract: AECO02

Lab Code: CHEM

Case No.: P4921

SAS No.: P4921

SDG NO.: P4921

Lab Sample ID: P4892-03MS

Date(s) Analyzed: 11/21/2024

11/21/2024

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR2

ID: 0.32 (mm)

GC Column: (2): ZB-MR1

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.50	7.45	7.55	4.70	13.9
	2	6.61	6.56	6.66	5.40	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	5.00	7.7
	2	3.61	3.56	3.66	5.40	
Heptachlor	1	4.91	4.86	4.96	5.10	7.5
	2	3.95	3.90	4.00	5.50	
Heptachlor epoxide	1	5.68	5.63	5.73	5.00	11.3
	2	4.73	4.68	4.78	5.60	
Endrin	1	6.57	6.52	6.62	4.80	17.1
	2	5.64	5.59	5.69	5.70	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WB-310-BOTMSD

Contract: AECO02

Lab Code: CHEM

Case No.: P4921

SAS No.: P4921

SDG NO.: P4921

Lab Sample ID: P4892-03MSD

Date(s) Analyzed: 11/21/2024

11/21/2024

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR2

ID: 0.32 (mm)

GC Column:(2): ZB-MR1

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.50	7.45	7.55	4.60	14.1
	2	6.61	6.56	6.66	5.30	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	5.00	5.8
	2	3.61	3.56	3.66	5.30	
Heptachlor	1	4.91	4.86	4.96	5.10	5.7
	2	3.95	3.90	4.00	5.40	
Heptachlor epoxide	1	5.68	5.63	5.73	5.00	9.5
	2	4.73	4.68	4.78	5.50	
Endrin	1	6.57	6.52	6.62	4.80	17.1
	2	5.64	5.59	5.69	5.70	