

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/21/2024

Lab Sample No.: PSTDCCC050 Data File : PL093193.D Time Analyzed: 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.057	8.954	9.154	45.380	50.000	-9.2
Endrin	6.572	6.474	6.674	43.030	50.000	-13.9
gamma-BHC (Lindane)	4.326	4.228	4.428	49.920	50.000	-0.2
Heptachlor	4.914	4.817	5.017	47.490	50.000	-5.0
Heptachlor epoxide	5.683	5.584	5.784	47.140	50.000	-5.7
Methoxychlor	7.500	7.399	7.599	41.800	50.000	-16.4
Tetrachloro-m-xylene	3.537	3.440	3.640	50.810	50.000	1.6

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/21/2024

Lab Sample No.: PSTDCCC050 Data File : PL093193.D Time Analyzed: 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.915	7.816	8.016	50.970	50.000	1.9
Endrin	5.638	5.541	5.741	50.120	50.000	0.2
gamma-BHC (Lindane)	3.607	3.511	3.711	52.430	50.000	4.9
Heptachlor	3.946	3.849	4.049	50.860	50.000	1.7
Heptachlor epoxide	4.728	4.632	4.832	52.370	50.000	4.7
Methoxychlor	6.613	6.515	6.715	47.990	50.000	-4.0
Tetrachloro-m-xylene	2.774	2.678	2.878	51.270	50.000	2.5