

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/21/2024

Lab Sample No.: PSTDCCC050 Data File : PL093211.D Time Analyzed: 15:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.058	8.954	9.154	46.110	50.000	-7.8
Endrin	6.572	6.474	6.674	44.290	50.000	-11.4
gamma-BHC (Lindane)	4.325	4.228	4.428	50.530	50.000	1.1
Heptachlor	4.914	4.817	5.017	47.570	50.000	-4.9
Heptachlor epoxide	5.682	5.584	5.784	47.490	50.000	-5.0
Methoxychlor	7.500	7.399	7.599	42.470	50.000	-15.1
Tetrachloro-m-xylene	3.536	3.440	3.640	51.100	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/21/2024

Lab Sample No.: PSTDCCC050 Data File : PL093211.D Time Analyzed: 15:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.914	7.816	8.016	51.630	50.000	3.3
Endrin	5.638	5.541	5.741	50.460	50.000	0.9
gamma-BHC (Lindane)	3.607	3.511	3.711	52.440	50.000	4.9
Heptachlor	3.946	3.849	4.049	50.770	50.000	1.5
Heptachlor epoxide	4.728	4.632	4.832	52.850	50.000	5.7
Methoxychlor	6.612	6.515	6.715	48.280	50.000	-3.4
Tetrachloro-m-xylene	2.774	2.678	2.878	51.250	50.000	2.5