

Surrogate Summary

SDG No.: **P4921**

Client: **AECOM**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL092652.D	PIBLK-PL092652.D	Decachlorobiphenyl	1	20	22.7	114		43	140
		Tetrachloro-m-xylene	1	20	21.6	108		77	126
		Decachlorobiphenyl	2	20	21.7	109		43	140
		Tetrachloro-m-xylene	2	20	20.4	102		77	126
I.BLK-PL093191.D	PIBLK-PL093191.D	Decachlorobiphenyl	1	20	19.6	98		43	140
		Tetrachloro-m-xylene	1	20	21.9	110		77	126
		Decachlorobiphenyl	2	20	21.3	106		43	140
		Tetrachloro-m-xylene	2	20	21.0	105		77	126
PB165164BL	PB165164BL	Decachlorobiphenyl	1	20	19.2	96		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.9	110		43	140
		Tetrachloro-m-xylene	2	20	19.5	98		77	126
PB165164BS	PB165164BS	Decachlorobiphenyl	1	20	19.9	99		43	140
		Tetrachloro-m-xylene	1	20	21.0	105		77	126
		Decachlorobiphenyl	2	20	21.7	109		43	140
		Tetrachloro-m-xylene	2	20	19.1	96		77	126
PB165123TB	PB165123TB	Decachlorobiphenyl	1	20	20.4	102		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	22.3	111		43	140
		Tetrachloro-m-xylene	2	20	19.8	99		77	126
P4892-03MS	WB-310-BOTMS	Decachlorobiphenyl	1	20	18.3	92		43	140
		Tetrachloro-m-xylene	1	20	20.3	102		77	126
		Decachlorobiphenyl	2	20	20.6	103		43	140
		Tetrachloro-m-xylene	2	20	18.4	92		77	126
P4892-03MSD	WB-310-BOTMSD	Decachlorobiphenyl	1	20	18.4	92		43	140
		Tetrachloro-m-xylene	1	20	19.8	99		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	18.1	90		77	126
P4921-01	WC-11-A-202411	Decachlorobiphenyl	1	20	17.4	87		43	140
		Tetrachloro-m-xylene	1	20	20.5	102		77	126
		Decachlorobiphenyl	2	20	19.4	97		43	140
		Tetrachloro-m-xylene	2	20	20.3	101		77	126
I.BLK-PL093210.D	PIBLK-PL093210.D	Decachlorobiphenyl	1	20	20.5	103		43	140
		Tetrachloro-m-xylene	1	20	21.9	109		77	126
		Decachlorobiphenyl	2	20	22.8	114		43	140
		Tetrachloro-m-xylene	2	20	21.0	105		77	126