

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4921

Client: AECOM

Analytical Method: 8081B

DataFile : PL093201.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	WB-310-BOTMSD											
P4892-03MSD	gamma-BHC (Lindane)	5	0	5.30	ug/L	106		2		60	152	20
	Heptachlor	5	0	5.40	ug/L	108		2		56	147	20
	Heptachlor epoxide	5	0	5.50	ug/L	110		2		77	143	20
	Endrin	5	0	5.70	ug/L	114		0		76	144	20
	Methoxychlor	5	0	5.30	ug/L	106		2		70	142	20