

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO02
 Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG No.: P4921
 Instrument ID: BNA_F Calibration Date/Time: 11/25/2024 09:33
 Lab File ID: BF140590.D Init. Calib. Date(s): 11/21/2024 11/21/2024
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:13 14:18
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.084	1.124		3.7	
2-Fluorophenol	1.172	1.119		-4.5	
Phenol-d6	1.550	1.507		-2.8	
1,4-Dichlorobenzene	1.435	1.399		-2.5	20.0
2-Methylphenol	1.010	0.988		-2.2	
3+4-Methylphenols	1.298	1.252		-3.5	
Nitrobenzene-d5	0.391	0.377		-3.6	
Hexachloroethane	0.536	0.520		-3.0	
Nitrobenzene	0.404	0.386		-4.5	
Hexachlorobutadiene	0.215	0.212		-1.4	20.0
2,4,6-Trichlorophenol	0.367	0.368		0.3	20.0
2-Fluorobiphenyl	1.342	1.315		-2.0	
2,4,5-Trichlorophenol	0.399	0.405		1.5	
2,4-Dinitrotoluene	0.397	0.407		2.5	
2,4,6-Tribromophenol	0.214	0.210		-1.9	
Hexachlorobenzene	0.238	0.233		-2.1	
Pentachlorophenol	0.105	0.110		4.8	20.0
Terphenyl-d14	1.284	1.204		-6.2	

All other compounds must meet a minimum RRF of 0.010.