

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4921

Client: AECOM

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4892-03MSD	Client Sample ID:	WB-310-BOTMSD					DataFile:	BF140614.D		
Pyridine	500	0	370	ug/L	74		8		10	109	20
1,4-Dichlorobenzene	500	0	290	ug/L	58		10		55	125	20
2-Methylphenol	500	0	430	ug/L	86		9		37	126	20
3+4-Methylphenols	500	0	440	ug/L	88		7		31	127	20
Hexachloroethane	500	0	280	ug/L	56		7		49	110	20
Nitrobenzene	500	0	400	ug/L	80		10		62	112	20
Hexachlorobutadiene	500	0	350	ug/L	70		8		52	125	20
2,4,6-Trichlorophenol	500	0	480	ug/L	96		8		78	112	20
2,4,5-Trichlorophenol	500	0	480	ug/L	96		8		71	111	20
2,4-Dinitrotoluene	500	0	500	ug/L	100		8		50	142	20
Hexachlorobenzene	500	0	460	ug/L	92		10		72	115	20
Pentachlorophenol	1000	0	1100	ug/L	110		9		25	139	20