

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/21/2024

Lab File ID: BF140539.D Time Analyzed: 16:54

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	106707	6.875	411509	8.16	234805	9.91
UPPER LIMIT	213414	7.375	823018	8.657	469610	10.41
LOWER LIMIT	53353.5	6.375	205755	7.657	117403	9.41
EPA SAMPLE NO.						
01 WC-11-A-202411	79300	6.87	290595	8.15	151253	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.