

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/27/2024

Lab File ID: BF140655.D Time Analyzed: 08:47

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	98372	6.869	362566	8.15	198240	9.91
UPPER LIMIT	196744	7.369	725132	8.651	396480	10.41
LOWER LIMIT	49186	6.369	181283	7.651	99120	9.41
EPA SAMPLE NO.						
01 PB165144BS	82046	6.87	308441	8.15	177190	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.