

### Surrogate Summary

SDG No.: P4921

Client: AECOM

Analytical Method: SW8260D

| Lab Sample ID | Client ID      | Parameter             | Spike | Result | RecoveryQual | Limits |      |
|---------------|----------------|-----------------------|-------|--------|--------------|--------|------|
|               |                |                       |       |        |              | Low    | High |
| P4921-01      | WC-11-A-202411 | 1,2-Dichloroethane-d4 | 50    | 49.2   | 98           | 74     | 125  |
|               |                | Dibromofluoromethane  | 50    | 46.0   | 92           | 75     | 124  |
|               |                | Toluene-d8            | 50    | 48.9   | 98           | 86     | 113  |
|               |                | 4-Bromofluorobenzene  | 50    | 48.0   | 96           | 77     | 121  |

## Surrogate Summary

SDG No.: P4921

Client: AECOM

Analytical Method: SW8260-Low

| Lab Sample ID | Client ID    | Parameter             | Spike | Result | RecoveryQual | Limits |      |
|---------------|--------------|-----------------------|-------|--------|--------------|--------|------|
|               |              |                       |       |        |              | Low    | High |
| VX1121WBL01   | VX1121WBL01  | 1,2-Dichloroethane-d4 | 50    | 49.5   | 99           | 74     | 125  |
|               |              | Dibromofluoromethane  | 50    | 46.0   | 92           | 75     | 124  |
|               |              | Toluene-d8            | 50    | 49.5   | 99           | 86     | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 49.0   | 98           | 77     | 121  |
| VX1121WBS01   | VX1121WBS01  | 1,2-Dichloroethane-d4 | 50    | 50.8   | 102          | 74     | 125  |
|               |              | Dibromofluoromethane  | 50    | 47.8   | 96           | 75     | 124  |
|               |              | Toluene-d8            | 50    | 48.3   | 97           | 86     | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 48.5   | 97           | 77     | 121  |
| VX1121WBSD0   | VX1121WBSD01 | 1,2-Dichloroethane-d4 | 50    | 49.4   | 99           | 74     | 125  |
|               |              | Dibromofluoromethane  | 50    | 46.4   | 93           | 75     | 124  |
|               |              | Toluene-d8            | 50    | 48.4   | 97           | 86     | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 49.5   | 99           | 77     | 121  |