

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4921

Client: AECOM

Analytical Method: SW8260-Low

Datafile : VX043935.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1121WBS01	Vinyl chloride	20	18.5	ug/L	93			65	117	
	1,1-Dichloroethene	20	17.9	ug/L	90			74	110	
	2-Butanone	100	95.6	ug/L	96			65	122	
	Carbon Tetrachloride	20	17.0	ug/L	85			77	113	
	Chloroform	20	19.1	ug/L	96			79	113	
	Benzene	20	18.4	ug/L	92			82	109	
	1,2-Dichloroethane	20	18.4	ug/L	92			80	115	
	Trichloroethene	20	15.5	ug/L	78			77	113	
	Tetrachloroethene	20	17.7	ug/L	89			67	123	
	Chlorobenzene	20	17.6	ug/L	88			82	109	