

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4921

Client: AECOM

Analytical Method: SW8260-Low

Datafile : VX043936.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1121WBSD01	Vinyl chloride	20	17.7	ug/L	89	4		65	117	20
	1,1-Dichloroethene	20	17.8	ug/L	89	1		74	110	20
	2-Butanone	100	98.3	ug/L	98	2		65	122	20
	Carbon Tetrachloride	20	17.8	ug/L	89	5		77	113	20
	Chloroform	20	18.3	ug/L	92	4		79	113	20
	Benzene	20	18.1	ug/L	91	1		82	109	20
	1,2-Dichloroethane	20	18.3	ug/L	92	0		80	115	20
	Trichloroethene	20	15.5	ug/L	78	0		77	113	20
	Tetrachloroethene	20	18.2	ug/L	91	2		67	123	20
	Chlorobenzene	20	17.5	ug/L	88	0		82	109	20