

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO02
 Lab Code: CHEM Case No.: P4921 SAS No.: P4921 SDG NO.: P4921
 Lab File ID: VX043928.D Date Analyzed: 11/21/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:17
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	100316	12.024				
UPPER LIMIT	200632	12.524				
LOWER LIMIT	50158	11.524				
EPA SAMPLE NO.						
WC-11-A-202411	87227	12.02				
VX1121WBL01	92124	12.02				
VX1121WBS01	103776	12.02				
VX1121WBSD01	93190	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.