

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4934 SAS No.: P4934 SDG No.: P4934

Instrument ID: BNA_N Calibration Date/Time: 11/21/2024 13:22

Lab File ID: BN035219.D Init. Calib. Date(s): 11/13/2024 11/13/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:40 16:15

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.713	0.604		-15.3	20.0
Fluoranthene-d10	1.225	0.956		-22.0	20.0
2-Fluorophenol	1.015	1.174		15.7	20.0
Phenol-d6	1.273	1.472		15.6	20.0
Nitrobenzene-d5	0.348	0.208		-40.2	20.0
2-Fluorobiphenyl	1.624	0.299		-81.6	20.0
2,4,6-Tribromophenol	0.289	0.331		14.5	20.0
Terphenyl-d14	0.839	3.629		332.5	20.0
1,4-Dioxane	0.363	0.403		11.0	20.0

All other compounds must meet a minimum RRF of 0.010.