

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4934 SAS No.: P4934 SDG No.: P4934

Instrument ID: BNA_N Calibration Date/Time: 11/25/2024 19:20

Lab File ID: BN035291.D Init. Calib. Date(s): 11/25/2024 11/25/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 10:16 13:52

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.560	0.573		2.3	50.0
Fluoranthene-d10	1.070	0.869		-18.8	50.0
2-Fluorophenol	0.756	0.966		27.8	50.0
Phenol-d6	1.056	1.244		17.8	50.0
Nitrobenzene-d5	0.324	0.194		-40.1	50.0
2-Fluorobiphenyl	0.221	0.144		-34.8	50.0
2,4,6-Tribromophenol	0.204	0.172		-15.7	50.0
Terphenyl-d14	7.392	8.425		14.0	50.0
1,4-Dioxane	0.288	0.380		31.9	50.0

All other compounds must meet a minimum RRF of 0.010.