

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	11/18/24
Project:	CTO WE13		Date Received:	11/20/24
Client Sample ID:	DUP01-20241118		SDG No.:	P4934
Lab Sample ID:	P4934-04		Matrix:	Water
Analytical Method:	SW8270SIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035229.D	1	11/21/24 08:27	11/21/24 19:24	PB165150

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.28		30 - 150		69%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		83%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.18	*	55 - 111		45%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.061	*	53 - 106		15%	SPK: 0.4
1718-51-0	Terphenyl-d14	2.81	*	58 - 132		702%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2750	7.322				
1146-65-2	Naphthalene-d8	6690	10.073				
15067-26-2	Acenaphthene-d10	4330	13.976				
1517-22-2	Phenanthrene-d10	8620	16.746				
1719-03-5	Chrysene-d12	1170	21.008				
1520-96-3	Perylene-d12	5970	23.095				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products