

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: P4934 SDG NO.: P4934

Lab File ID: BN035276.D

DFTPP Injection Date: 11/25/2024

Instrument ID: BNA_N

DFTPP Injection Time: 09:37

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.2
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	30.5
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	40.9
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	28.2
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	10.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	12.3 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN035277.D	11/25/2024	10:16
SSTDICC0.2	SSTDICC0.2	BN035278.D	11/25/2024	10:52
SSTDICCC0.4	SSTDICCC0.4	BN035279.D	11/25/2024	11:28
SSTDICC0.8	SSTDICC0.8	BN035280.D	11/25/2024	12:04
SSTDICC1.6	SSTDICC1.6	BN035281.D	11/25/2024	12:40
SSTDICC3.2	SSTDICC3.2	BN035282.D	11/25/2024	13:16
SSTDICC5.0	SSTDICC5.0	BN035283.D	11/25/2024	13:52
PB165150BL	PB165150BL	BN035285.D	11/25/2024	15:44
BPOW6-11-20241118	P4934-02	BN035286.D	11/25/2024	16:20
BPOW6-10-20241119	P4934-05	BN035287.D	11/25/2024	16:56
BPOW6-10-20241119MS	P4934-06MS	BN035288.D	11/25/2024	17:32
BPOW6-10-20241119MSD	P4934-07MSD	BN035289.D	11/25/2024	18:08
PB165150BS	PB165150BS	BN035290.D	11/25/2024	18:44
SSTDCCC0.4EC	SSTDCCC0.4	BN035291.D	11/25/2024	19:20