

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4934 SAS No.: P4934 SDG No.: P4934

Instrument ID: MSVOA_X Calibration Date/Time: 11/21/2024 20:18

Lab File ID: VX043948.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.667	0.697	0.1	4.5	50
Vinyl Chloride	0.709	0.706		-0.42	50
Bromomethane	0.436	0.421		-3.44	50
Chloroethane	0.432	0.403		-6.71	50
Trichlorofluoromethane	1.268	1.142		-9.94	50
1,1,2-Trichlorotrifluoroethane	0.605	0.548		-9.42	50
1,1-Dichloroethene	0.576	0.550		-4.51	50
Acetone	0.394	0.312		-20.81	50
Carbon Disulfide	1.188	1.054		-11.28	50
Methyl tert-butyl Ether	2.092	2.128		1.72	50
Methylene Chloride	0.668	0.639		-4.34	50
trans-1,2-Dichloroethene	0.595	0.587		-1.35	50
1,1-Dichloroethane	1.161	1.175	0.1	1.21	50
2-Butanone	0.508	0.482		-5.12	50
Carbon Tetrachloride	0.500	0.458		-8.4	50
cis-1,2-Dichloroethene	0.735	0.740		0.68	50
Chloroform	1.249	1.251		0.16	50
1,1,1-Trichloroethane	1.077	1.078		0.09	50
Methylcyclohexane	0.578	0.507		-12.28	50
Benzene	1.387	1.328		-4.25	50
1,2-Dichloroethane	0.557	0.528		-5.21	50
Trichloroethene	0.393	0.327		-16.79	50
1,2-Dichloropropane	0.340	0.333		-2.06	50
Bromodichloromethane	0.482	0.468		-2.9	50
4-Methyl-2-Pentanone	0.537	0.526		-2.05	50
Toluene	0.830	0.807		-2.77	50
t-1,3-Dichloropropene	0.491	0.468		-4.68	50
cis-1,3-Dichloropropene	0.539	0.513		-4.82	50
1,1,2-Trichloroethane	0.343	0.325		-5.25	50
2-Hexanone	0.403	0.387		-3.97	50
Dibromochloromethane	0.344	0.326		-5.23	50
Tetrachloroethene	0.330	0.303		-8.18	50
Chlorobenzene	1.098	1.042	0.3	-5.1	50
Ethyl Benzene	1.861	1.802		-3.17	50
m/p-Xylenes	0.694	0.662		-4.61	50
o-Xylene	0.678	0.664		-2.07	50
Styrene	1.121	1.112		-0.8	50
Bromoform	0.240	0.231	0.1	-3.75	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4934 SAS No.: P4934 SDG No.: P4934

Instrument ID: MSVOA_X Calibration Date/Time: 11/21/2024 20:18

Lab File ID: VX043948.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.843	3.715		-3.33	50
1,1,2,2-Tetrachloroethane	1.316	1.276	0.3	-3.04	50
1,3-Dichlorobenzene	1.670	1.554		-6.95	50
1,4-Dichlorobenzene	1.702	1.549		-8.99	50
1,2-Dichlorobenzene	1.685	1.619		-3.92	50
1,2-Dichloroethane-d4	0.858	0.898		4.66	50
Dibromofluoromethane	0.357	0.357		0	50
Toluene-d8	1.232	1.228		-0.32	50
4-Bromofluorobenzene	0.419	0.423		0.95	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.