

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4934 SAS No.: P4934 SDG No.: P4934

Instrument ID: MSVOA\_X Calibration Date/Time: 11/22/2024 21:05

Lab File ID: VX043975.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Chloromethane                  | 0.667 | 0.604  | 0.1        | -9.44  | 50    |
| Vinyl Chloride                 | 0.709 | 0.633  |            | -10.72 | 50    |
| Bromomethane                   | 0.436 | 0.468  |            | 7.34   | 50    |
| Chloroethane                   | 0.432 | 0.412  |            | -4.63  | 50    |
| Trichlorofluoromethane         | 1.268 | 1.357  |            | 7.02   | 50    |
| 1,1,2-Trichlorotrifluoroethane | 0.605 | 0.544  |            | -10.08 | 50    |
| 1,1-Dichloroethene             | 0.576 | 0.528  |            | -8.33  | 50    |
| Acetone                        | 0.394 | 0.292  |            | -25.89 | 50    |
| Carbon Disulfide               | 1.188 | 1.034  |            | -12.96 | 50    |
| Methyl tert-butyl Ether        | 2.092 | 2.090  |            | -0.1   | 50    |
| Methylene Chloride             | 0.668 | 0.624  |            | -6.59  | 50    |
| trans-1,2-Dichloroethene       | 0.595 | 0.569  |            | -4.37  | 50    |
| 1,1-Dichloroethane             | 1.161 | 1.155  | 0.1        | -0.52  | 50    |
| 2-Butanone                     | 0.508 | 0.453  |            | -10.83 | 50    |
| Carbon Tetrachloride           | 0.500 | 0.488  |            | -2.4   | 50    |
| cis-1,2-Dichloroethene         | 0.735 | 0.728  |            | -0.95  | 50    |
| Chloroform                     | 1.249 | 1.246  |            | -0.24  | 50    |
| 1,1,1-Trichloroethane          | 1.077 | 1.102  |            | 2.32   | 50    |
| Methylcyclohexane              | 0.578 | 0.512  |            | -11.42 | 50    |
| Benzene                        | 1.387 | 1.324  |            | -4.54  | 50    |
| 1,2-Dichloroethane             | 0.557 | 0.541  |            | -2.87  | 50    |
| Trichloroethene                | 0.393 | 0.317  |            | -19.34 | 50    |
| 1,2-Dichloropropane            | 0.340 | 0.329  |            | -3.23  | 50    |
| Bromodichloromethane           | 0.482 | 0.474  |            | -1.66  | 50    |
| 4-Methyl-2-Pentanone           | 0.537 | 0.509  |            | -5.21  | 50    |
| Toluene                        | 0.830 | 0.802  |            | -3.37  | 50    |
| t-1,3-Dichloropropene          | 0.491 | 0.480  |            | -2.24  | 50    |
| cis-1,3-Dichloropropene        | 0.539 | 0.525  |            | -2.6   | 50    |
| 1,1,2-Trichloroethane          | 0.343 | 0.324  |            | -5.54  | 50    |
| 2-Hexanone                     | 0.403 | 0.376  |            | -6.7   | 50    |
| Dibromochloromethane           | 0.344 | 0.327  |            | -4.94  | 50    |
| Tetrachloroethene              | 0.330 | 0.305  |            | -7.58  | 50    |
| Chlorobenzene                  | 1.098 | 1.047  | 0.3        | -4.64  | 50    |
| Ethyl Benzene                  | 1.861 | 1.812  |            | -2.63  | 50    |
| m/p-Xylenes                    | 0.694 | 0.689  |            | -0.72  | 50    |
| o-Xylene                       | 0.678 | 0.676  |            | -0.29  | 50    |
| Styrene                        | 1.121 | 1.146  |            | 2.23   | 50    |
| Bromoform                      | 0.240 | 0.239  | 0.1        | -0.42  | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4934 SAS No.: P4934 SDG No.: P4934

Instrument ID: MSVOA\_X Calibration Date/Time: 11/22/2024 21:05

Lab File ID: VX043975.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                  | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|---------------------------|-------|--------|------------|-------|-------|
| Isopropylbenzene          | 3.843 | 3.785  |            | -1.51 | 50    |
| 1,1,2,2-Tetrachloroethane | 1.316 | 1.298  | 0.3        | -1.37 | 50    |
| 1,3-Dichlorobenzene       | 1.670 | 1.571  |            | -5.93 | 50    |
| 1,4-Dichlorobenzene       | 1.702 | 1.595  |            | -6.29 | 50    |
| 1,2-Dichlorobenzene       | 1.685 | 1.612  |            | -4.33 | 50    |
| 1,2-Dichloroethane-d4     | 0.858 | 0.889  |            | 3.61  | 50    |
| Dibromofluoromethane      | 0.357 | 0.358  |            | 0.28  | 50    |
| Toluene-d8                | 1.232 | 1.216  |            | -1.3  | 50    |
| 4-Bromofluorobenzene      | 0.419 | 0.417  |            | -0.48 | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.