

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4934

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX043935.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1121WBS01	Chloromethane	20	19.4	ug/L	97			50	139	
	Vinyl chloride	20	18.5	ug/L	93			58	137	
	Bromomethane	20	18.7	ug/L	94			53	141	
	Chloroethane	20	20.7	ug/L	104			60	138	
	Trichlorofluoromethane	20	17.3	ug/L	86			65	141	
	1,1,2-Trichlorotrifluoroethane	20	17.6	ug/L	88			70	136	
	1,1-Dichloroethene	20	17.9	ug/L	90			71	131	
	Acetone	100	90.6	ug/L	91			39	160	
	Carbon disulfide	20	17.2	ug/L	86			64	133	
	Methyl tert-butyl Ether	20	19.4	ug/L	97			71	124	
	Methylene Chloride	20	18.6	ug/L	93			74	124	
	trans-1,2-Dichloroethene	20	18.5	ug/L	93			75	124	
	1,1-Dichloroethane	20	19.7	ug/L	99			77	125	
	2-Butanone	100	95.6	ug/L	96			56	143	
	Carbon Tetrachloride	20	17.0	ug/L	85			72	136	
	cis-1,2-Dichloroethene	20	18.9	ug/L	95			78	123	
	Chloroform	20	19.1	ug/L	96			79	124	
	1,1,1-Trichloroethane	20	18.9	ug/L	95			74	131	
	Methylcyclohexane	20	16.9	ug/L	85			72	132	
	Benzene	20	18.4	ug/L	92			79	120	
	1,2-Dichloroethane	20	18.4	ug/L	92			73	128	
	Trichloroethene	20	15.5	ug/L	78		*	79	123	
	1,2-Dichloropropane	20	18.6	ug/L	93			78	122	
	Bromodichloromethane	20	18.1	ug/L	91			79	125	
	4-Methyl-2-Pentanone	100	93.9	ug/L	94			67	130	
	Toluene	20	18.9	ug/L	95			80	121	
	t-1,3-Dichloropropene	20	18.1	ug/L	91			73	127	
	cis-1,3-Dichloropropene	20	18.3	ug/L	92			75	124	
	1,1,2-Trichloroethane	20	18.2	ug/L	91			80	119	
	2-Hexanone	100	93.9	ug/L	94			57	139	
	Dibromochloromethane	20	17.5	ug/L	88			74	126	
	Tetrachloroethene	20	17.7	ug/L	89			74	129	
	Chlorobenzene	20	17.6	ug/L	88			82	118	
	Ethyl Benzene	20	18.4	ug/L	92			79	121	
	m/p-Xylenes	40	37.3	ug/L	93			80	121	
	o-Xylene	20	18.8	ug/L	94			78	122	
	Styrene	20	18.8	ug/L	94			78	123	
	Bromoform	20	16.6	ug/L	83			66	130	
	Isopropylbenzene	20	18.9	ug/L	95			72	131	
	1,1,2,2-Tetrachloroethane	20	18.9	ug/L	95			71	121	
	1,3-Dichlorobenzene	20	17.9	ug/L	90			80	119	
	1,4-Dichlorobenzene	20	17.7	ug/L	89			79	118	
	1,2-Dichlorobenzene	20	18.0	ug/L	90			80	119	