

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4934 SAS No.: P4934 SDG NO.: P4934
Lab File ID: VX043928.D Date Analyzed: 11/21/2024
Instrument ID: MSVOA_X Time Analyzed: 11:17
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	138578	5.54	241629	6.76	212243	10.05
UPPER LIMIT	277156	6.038	483258	7.257	424486	10.549
LOWER LIMIT	69289	5.038	120815	6.257	106122	9.549
EPA SAMPLE NO.						
TB01-20241118	113495	5.55	211780	6.76	190080	10.06
BPOW6-11-20241118	120827	5.54	229101	6.76	190728	10.06
BPOW6-7-20241118	120589	5.55	229817	6.76	203544	10.06
DUP01-20241118	120506	5.54	229149	6.76	199026	10.06
BPOW6-10-20241119	110999	5.54	211108	6.76	188768	10.06
BPOW6-9-20241119	110330	5.55	206208	6.76	181398	10.06
FB01-20241119	112214	5.54	209177	6.76	184189	10.06
BPOW6-8-20241119	120427	5.55	226777	6.76	198351	10.06
RB01-20241119	113782	5.55	213883	6.76	191929	10.06
DUP02-20241119	108515	5.55	209984	6.76	184320	10.06
VX1121WBL01	130697	5.54	247747	6.76	216577	10.05
VX1121WBS01	142766	5.54	267012	6.76	228045	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.