



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

RunNo.: LB133556

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.498	0.50	100	90-110	11/21/2024
Sample ID: CCV1 Phosphorus, Total	mg/L	0.497	0.50	99	90-110	11/21/2024
Sample ID: CCV2 Phosphorus, Total	mg/L	0.498	0.50	100	90-110	11/21/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

RunNo.: LB133558

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.500	0.50	100	90-110	11/21/2024
Sample ID: CCV1 Orthophosphate as P	mg/L	0.500	0.5	100	90-110	11/21/2024
Sample ID: CCV2 Orthophosphate as P	mg/L	0.500	0.5	100	90-110	11/21/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

RunNo.: LB133582

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Ammonia as N		mg/L	1	1	100	90-110	11/22/2024
Sample ID:	CCV1						
Ammonia as N		mg/L	0.98	1	98	90-110	11/22/2024
Sample ID:	CCV2						
Ammonia as N		mg/L	1	1	100	90-110	11/22/2024
Sample ID:	CCV3						
Ammonia as N		mg/L	1	1	100	90-110	11/22/2024