

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: P4937
Project: Pre Treatment Plant	Sample ID: P4900-06
Client ID: C0K69MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	20.8		20.3		1	2.43		11/21/2024

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	P4937
Project:	Pre Treatment Plant	Sample ID:	P4931-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	425		443		1	4.15		11/21/2024

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	P4937
Project:	Pre Treatment Plant	Sample ID:	P4937-01
Client ID:	EFFLUENTDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	0.069		0.068		1	1.46		11/21/2024
Orthophosphate as P	mg/L	+/-20	0.057		0.057		1	0		11/21/2024

Duplicate Sample Summary

Client: Holland Manufacturing Co. Project: Pre Treatment Plant Client ID: EFFLUENTMSD	SDG No.: P4937 Sample ID: P4937-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	0.55		0.55		1	0.55		11/21/2024
Orthophosphate as P	mg/L	+/-20	0.54		0.54		1	0.37		11/21/2024

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	P4937
Project:	Pre Treatment Plant	Sample ID:	P4937-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	72.8		72.4		1	0.55		11/21/2024

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	P4937
Project:	Pre Treatment Plant	Sample ID:	P4947-01
Client ID:	A3988DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	0.46		0.46		1	0		11/22/2024

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	P4937
Project:	Pre Treatment Plant	Sample ID:	P4947-01
Client ID:	A3988MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	1.40		1.40		1	0		11/22/2024

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: P4937
Project: Pre Treatment Plant	Sample ID: P4967-01
Client ID: 001-WILLETS-PT-BLVDDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	38.8		40.0		1	3.05		11/25/2024